



THE GLOBAL STANDARD
FOR LIVESTOCK DATA

Network. Guidelines. Certification.

FRAME OF ACTIVITY : ICAR MILK ANALYSES
SUB-COMMITTEE (MA SC)

ICAR - Proficiency Testing Scheme Raw cow milk

“Routine” Methods

ORGANISER: ICAR, VIA SAVOIA 78, I-00198 ROME, ITALY






Table of contents

1. Introduction
2. Your performances analysis
3. Control Charts
4. ICAR Statistical Elaboration
 - Fat (routine method)
 - Protein (routine method)
 - Lactose (routine method)
 - Urea (routine method)
 - BHB (chemical and routine method)
 - PAG (ELISA method)



1.Introduction

Dear Participant,

Thank you for participating in the ICAR Proficiency Test (PT) March 2017 !

In this report you will find sections 2 and 3 which are dedicated to “your” quality assurance management and section 4 dedicated to the “general” statistical elaboration for each parameter.

The proficiency test is a tool to help evaluate the performance of the laboratory process and to support your laboratory quality assurance system. Its aim is to provide independent data for you to monitor, evaluate and ultimately improve your processes as you see fit.

From the analyses of the data received we have identified some aspects that if evaluated and managed may serve to improve some control steps of your quality management ISO 17025.

When the PT samples arrive to your laboratory they can be viewed as being from a ‘customer’ that is asking you to provide timely, precise and accurate results.

In tables A,B,C,D,E,F,G if all the information is reported correctly from the participant, then the cells are filled in green, otherwise they are highlighted in red for your attention, so you can review and verify any causal reasons internally. The control charts H will help you to follow your performance over the time.

- A) In table A you find your participation codes and the information if all the results from the samples received, have been sent to the PT provider.
- B) In table B is indicated if the results have been sent on time.
- C) In table C is indicated if the results have been reported in the correct unit of measurements.
- D) In this table is reported if some of your results are outlier for Grubbs or Cochran test
- E) The evaluation of repeatability of the results should be one of the first controls before communication of the data. In table E the absolute difference between replicates is compared with the repeatability limit of the relevant “reference” method indicated. If one or more results have a result out of the limit, the cell is in red. It may be that you have deployed a chemical method that is different from the reference method indicated. If the repeatability is bigger it will be evaluated internally with the precision of the specific method used. You can find all the detailed information of your data in Table II in the section Statistical elaboration for each parameter.
- F) In table E the results of your $Z\text{-Score}_{PT}$ (standard deviation calculated on this proficiency test) and the $Z\text{-Score}_{FIX}$ (standard deviation of the reference method) are summarized. If you have obtained all the $-2 < Z\text{-Score results} < +2$ the cell will be filled in green. If you have obtained one or more results in the moderate or poor performance range the cells will be filled in yellow or red respectively.
- G) In table F the ranking of your lab will be green if the mean of difference and standard deviation of difference value are in the box of figure 2 of each parameter. Limits are only

ICAR Proficiency Test (PT) “Routine” methods – March 2017



indicative and so far do not constitute standard values; they indicate what is normally reachable by labs for their self evaluation. According the results obtained the MA SC will decide eventually to revise.

- H) The control charts have been created using your data obtained with the $Z\text{-Score}_{PT}$ and $Z\text{-Score}_{FIX}$. The standard deviation of the method will be the same in the next PT round and it will facilitate the comparison of your performance over the time.

For the BHB we have not established a fix standard deviation so we are providing you only the ZS for the PT

Your Control Charts have been updated with data of the previous ICAR PT if you have participated.

The sample preparation and statistical elaboration have been done by ICAR Sub- contractor Actalia, accredited to ISO 17043.

In the second part of the report the statistical elaboration followed the template approved by ICAR's Milk Analyses Sub Committee chaired by Dr. Gavin Scott (NZ). You find the statistical elaboration for all the ICAR interested parameters, fat, protein, lactose and urea.

We think it is important to show you, as ICAR member, the reproducibility of the ICAR laboratories, even if you have not participated in this PT round.

For each parameter the SR=standard deviation of reproducibility has been calculated after the outlier elimination. If you have participated, and your results are in the repeatability limits, you can use this value for the calculation of your uncertainty of measurement.

ICAR would like to see, in the next years, this part 4 of this report, completed with the results, reference and/or routine methods, from all the ICAR countries for the parameters indicated.

We are sure with your support and contribution it will grow to benefit all!

The list of all ICAR reference laboratories and those participated in ICAR PT 2016 with at least one parameter is reported below and upload on ICAR website

<http://www.icar.org/index.php/certifications/milk-analysis-laboratories-certifications/rresults-of-the-milk-laboratories-proficiency-test-march-2017>

ICAR MILK PROFICIENCY TEST - MARCH 2017
COUNTRIES WITH AT LEAST ONE LABORATORY
PARTICIPATING TO THE PT (listed in dark)



Laboratory	Country
Eastern Lab services	USA
Qlip B.V.,	The Netherlands
Teagasc Food research Center	Ireland
Milchprüfring Baden-Württemberg e.V.	Germany
Laboratorio Standard Latte	Italy
Central Milk Lab ICBA	Israel
Cattle Information Service (CIS)	England
Eurofins Steins Laboratory A/S	Denmark
Valacta - Centre d'Expertise en Production Laitière du Québec	Canada
Estonian Livestock Performance Recording Ltd	Estonia

ICAR Proficiency Test (PT) "Routine" methods – March 2017

Network. Guidelines. Certification.



Laboratory	Country
Laboratorium Oceny Mleka KCHZ Laboratorium Referecyjne z/s w Parzniewie	Poland
Japan Dairy Technical Association	Japan
PFHBiPM Region Oceny Bydgoszcz z/s w Minikowie	Poland
Laboratorija za ispitivanje kvaliteta mleka, Poljoprivredni fakultet Novi Sad,	Serbia
University of Ljubljana Biotechnical faculty department of Animal Science Institute of Dairy Science and Probiotics	Slovenia
Federazione Latterie Alto Adige Soc. Agr. Coop.	Italy
Actalia	Finland
Direction de l' Amelioration Genetique Direction de Amelioration Genetique	Tunisia
Laborator pro rozbor mléka Bustehrad, Ceskomoravská společnost chovatelů a.s.	Czech Rep.
Merieux Nutriscience South Africa	South Africa
Valio	Finland
Central Milk Testing Lab	Canada
Pacific Milk Analysis	Canada
Horizon Lab Ltd	Canada
Croatian Agricultural Agency, Central Laboratory for Milk Quality Control	Croatia
Delaval international AB	Sweden
COOPRINSEM	Chile
Merieux NutriScience	South Africa
Holstein-Ro Breeders Association Laboratory	Romania
Eurofins Steins Laboratory A/B	Sweden
Shangai dairy breeding center Co.Ltd	China

ICAR would like to stay at your side to support you in any way we can to help improve overall quality management systems for milk analyses. Your active participation in the ICAR PTs and in the Milk Analyses meetings is encouraging. We welcome any and all feedback/comments you may have on this activity, as it will help us continuously improve and to ultimately provide you a better service.

Kind Regards,

ICAR Secretariat





Routine Methods

ICAR PT
September 2016

Laboratory participation codes and Performance analyses

Laboratory Name	Laboratory name
-----------------	-----------------

A	Your participation Codes				
	Subscription	Fat	Protein	Lactose	Urea
	Participation Codes	13	13		
	Are all the sample results received ?	Yes	Yes		

B	Data results received on time		
	Yes	No	Deadline
			19.09.2016

C	Have you sent the data with the correct units of measurements ?				
	Fat	Protein	Lactose	Urea	
	g/100g	g/100g	g/100g	mg/dl	
	Yes	Yes			

D	Outliers				
	Fat	Protein	Lactose	Urea	
	Sample	Sample	Sample	Sample	
	g/100g	g/100g	g/100g	mg/dl	

E	Repeatability				
	Your "r" performance				Limits
	Fat	Protein	Lactose	Urea	
	g/100g	g/100g	g/100g	mg/dl	

If the repeatability is smaller than the limit the cell is in green if there is a sample with a "r" bigger than the limit the cell is in red. Please check table II in correspondence of the parameter and your lab code.

F	Z-Score				
	Your Z-Score PT				Interpretation Z-Score
	Fat	Protein	Lactose	Urea	

If there is a sample with a "z-score" in the yellow or red area please check table VI and VII in correspondence of your lab code.

G	Ranking of your lab				
	Mean difference and standard deviation of difference				Indicative Limits defined in the ICAR MA SC
	Fat	Protein	Lactose	Urea	

If d and sd are in the limit (see Table 1 and Figure 1) the cells are in green



Note : Limits are only indicative and so far do not constitute standard values; they indicate what is normally reachable by labs for their self evaluation

Legenda:



	positive performance for all the sample
	same sample showed a performance out of range or there are miss data
	The parameter was not analyzed

Control Charts

Control Charts Fat



Control Charts Protein





ICAR
PROFICIENCY TESTING SCHEME

March 2017

Raw Milk

Determination of FAT CONTENT
Routine method

Repe
Repe

Sending date of statistical treatment : 21th April 2017

Frame of activity :	ICAR Milk Analyses Sub Committee (MA SC)		
Contact :	Gavin Scott	gavin@milktest.co.nz	
ICAR Staff	Silvia Orlandini	pt@icar.org	silvia@icar.org

Table I : Ranking of the laboratories

Units : g / 100 g

Nb	%	N°	d	Sd	D	Method
1	5	1	+ 0,003	0,006	0,007	IR
2	10	14	- 0,007	0,013	0,015	IR
3	14	15	+ 0,010	0,012	0,016	IR
4	19	8	+ 0,016	0,009	0,019	IR
5	24	5	- 0,005	0,018	0,019	IR
6	29	16	+ 0,020	0,008	0,022	IR
7	33	6	- 0,014	0,018	0,023	IR
8	38	17	- 0,009	0,025	0,027	IR
9	43	4	- 0,025	0,015	0,029	IR
10	48	2	+ 0,018	0,028	0,033	Other*
11	52	21	+ 0,038	0,011	0,039	IR
12	57	3	+ 0,038	0,013	0,040	IR
13	62	11	+ 0,010	0,046	0,047	IR
14	67	12	+ 0,045	0,028	0,053	IR
15	71	20	- 0,052	0,015	0,054	IR
16	76	10	- 0,016	0,053	0,055	IR
17	81	7	+ 0,031	0,048	0,057	IR
18	86	13	- 0,055	0,024	0,060	IR
19	90	9	- 0,032	0,072	0,079	IR
20	95	18	+ 0,059	0,122	0,135	IR
21	100	19	- 0,084	0,137	0,161	IR

The table should be studied in parallel with figure 1 where the laboratories are located according to an acceptability area (or target) the limits of which are :

+/- 0,020 g / 100 g for $\bar{d}$ and 0,030 g / 100 g for Sd

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 20 sets of results send by 20 laboratories using routine method ISO 9622 | IDF 141, after outlier discarding using Grubbs test at 5 % risk level

Other* : ASU L 01.00-78 IR-mesurement

(NC : OUT of RANKING because of insufficient data number)

(Nb : laboratory rank; % : relative rank)

(N° : laboratory identification number)

(d et Sd : mean and standard deviation of the differences (laboratory -reference))

(D : Euclidian distance to YX-axis origin = SQUARE ROOT.(d² + Sd²))

Note : Limits are only indicative and so far do not constitute standard values; they indicate what is normally reachable by labs for their self evaluation.

repeatability standard deviation of this ICAR proficiency test (after Cochran elimination at 5 %)

Sr_{PT} 0,005

reproducibility standard deviation of this ICAR proficiency test (after Cochran and Grubbs elimination at 5 %)

SR_{PT} 0,040

Table II : REPEATABILITY - Absolute difference between replicates in g / 100 g

Sample Lab Code	1	2	3	4	5	6	7	8	9	10	Sr	NL
1	0,010	0,010	0,000	0,010	0,000	0,010	0,010	0,010	0,010	0,010	0,006	20
2	0,010	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,002	20
3	0,010	0,000	0,020 *	0,020	0,010	0,010	0,010	0,010	0,010	0,010	0,009	20
4	0,010	0,010	0,000	0,000	0,000	0,010	0,000	0,000	0,000	0,010	0,004	20
5	0,000	0,000	0,000	0,010	0,000	0,010	0,010	0,000	0,000	0,000	0,004	20
6	0,010	0,000	0,010	0,010	0,000	0,010	0,000	0,000	0,000	0,000	0,004	20
7	0,002	0,017	0,005	0,005	0,020 *	0,010	0,010	0,010	0,012	0,014	0,008	20
8	0,000	0,000	0,000	0,000	0,000	0,010	0,000	0,000	0,000	0,000	0,002	20
9	0,000	0,000	0,000	0,000	0,000	0,010	0,000	0,000	0,010	0,000	0,003	20
10	0,010	0,020	0,010	0,010	0,010	0,020	0,010	0,010	0,010	0,010	0,009	20
11	0,010	0,010	0,000	0,000	0,010	0,020	0,010	0,000	0,010	0,010	0,007	20
12	0,030 *	0,000	0,010	0,010	0,000	0,010	0,000	0,000	0,010	0,000	0,008	20
13	0,000	0,000	0,000	0,010	0,000	0,000	0,010	0,000	0,000	0,010	0,004	20
14	0,010	0,010	0,005	0,005	0,000	0,005	0,000	0,005	0,005	0,005	0,004	20
15	0,010	0,000	0,000	0,010	0,010	0,000	0,030 *	0,010	0,000	0,000	0,008	20
16	0,010	0,010	0,010	0,000	0,000	0,010	0,000	0,010	0,000	0,000	0,005	20
17	0,010	0,000	0,000	0,000	0,000	0,000	0,010	0,010	0,000	0,000	0,004	20
18	0,010	0,000	0,000	0,010	0,010	0,000	0,010	0,000	0,000	0,010	0,005	20
19	0,040 *	0,010	0,000	0,010	0,000	0,050 *	0,020	0,000	0,010	0,030 *	0,017	20
20	0,000	0,000	0,000	0,010	0,000	0,000	0,010	0,000	0,000	0,000	0,003	20
21	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	20
Sr	0,009	0,006	0,004	0,006	0,005	0,010	0,007	0,004	0,005	0,007		420
NE	42	42	42	42	42	42	42	42	42	42		
L	0,022	0,022	0,013	0,023	0,014	0,027	0,023	0,017	0,018	0,019		

Sr : repeatability standard deviation of each laboratory limit 0,014 g/100g

NL : number of measurements per laboratory

L : Limit for difference between duplicates according Cochran test at 5% level.

SE : repeatability standard deviation per sample

NE : number of measurements per sample

*: discarded data using the test of Cochran at 5 %

** : missing data

r : limit of repeatability, absolute difference between two replicates=0,040 according ISO 9622|IDF 141

Table III : Means of the replicates in g / 100 g

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	4,805	4,195	3,510	2,805	2,120	4,535	3,855	3,105	2,425	1,455
2	4,845	4,240	3,540	2,810	2,110	4,580	3,880	3,120	2,410	1,430
3	4,825	4,220	3,540	2,840	2,175	4,565	3,875	3,135	2,475	1,515
4	4,765	4,165	3,490	2,790	2,100	4,475	3,810	3,090	2,410	1,435
5	4,780	4,180	3,500	2,805	2,140	4,505	3,815	3,100	2,430	1,480
6	4,755	4,170	3,495	2,805	2,130	4,495	3,830	3,100	2,420	1,440
7	4,929	4,222	3,518	2,810	2,120	4,645	3,845	3,135	2,404	1,465
8	4,810	4,200	3,520	2,820	2,150	4,545	3,850	3,120	2,440	1,490
9	4,750	4,140	3,440	2,740	2,050	4,545	3,850	3,250 *	2,345	1,350
10	4,815	4,200	3,515	2,805	2,115	4,370	3,845	3,105	2,415	1,435
11	4,775	4,175	3,500	2,800	2,125	4,660	3,815	3,090	2,455	1,485
12	4,865	4,230	3,565	2,775	2,170	4,605	3,900	3,150	2,445	1,530
13	4,710	4,120	3,450	2,755	2,090	4,450	3,785	3,050	2,380	1,445
14	4,815	4,195	3,503	2,788	2,110	4,548	3,820	3,088	2,408	1,438
15	4,835	4,200	3,520	2,825	2,125	4,550	3,845	3,105	2,420	1,460
16	4,825	4,205	3,525	2,820	2,140	4,565	3,860	3,115	2,440	1,490
17	4,795	4,200	3,540	2,790	2,100	4,540	3,845	3,095	2,400	1,390
18	5,205 *	4,190	3,560	2,815	2,145	4,560	3,835	3,130	2,440	1,495
19	4,750	4,225	3,540	2,745	2,140	4,415	3,790	3,120	2,095 *	1,125 *
20	4,730	4,120	3,460	2,755	2,080	4,470	3,785	3,050	2,380	1,430
21	4,860	4,230	3,550	2,830	2,150	4,590	3,870	3,140	2,450	1,490
M	4,802	4,192	3,513	2,797	2,123	4,534	3,838	3,107	2,420	1,457
REF.	4,800	4,189	3,512	2,796	2,124	4,532	3,836	3,106	2,420	1,459
SD	0,052	0,034	0,034	0,028	0,030	0,071	0,031	0,026	0,030	0,042

M = mean per sample

REF. = reference values

SD = standard deviation per sample

*: discarded data using the test of Grubbs at 5 %

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 20 laboratories using the Routine method ISO 9622 | IDF 141, after outliers discarding using Grubbs test at 5 % risk level.

Table IV : Outlier identification

Sample	1	2	3	4	5	6	7	8	9	10
Outliers Cochran	12; 19		3		7	19	15			19
Outlier Grubbs	18							9	19	19
sr	0,006	0,006	0,003	0,006	0,004	0,007	0,006	0,004	0,004	0,005
SR	0,052	0,034	0,034	0,028	0,030	0,067	0,032	0,027	0,030	0,042

Table V : ACCURACY - differences (laboratory - reference) in g / 100 g

Sample Lab Code	1	2	3	4	5	6	7	8	9	10	d	Sd _{lab}	t
1	+ 0,005	+ 0,006	- 0,002	+ 0,009	- 0,004	+ 0,003	+ 0,019	- 0,001	+ 0,005	- 0,004	+ 0,004	0,007	1,65
2	+ 0,045	+ 0,051	+ 0,028	+ 0,014	- 0,014	+ 0,048	+ 0,044	+ 0,014	- 0,010	- 0,029	+ 0,019	0,029	2,09
3	+ 0,025	+ 0,031	+ 0,028	+ 0,044	+ 0,051	+ 0,033	+ 0,039	+ 0,029	+ 0,055	+ 0,056	+ 0,039	0,012	10,55
4	- 0,035	- 0,024	- 0,022	- 0,006	- 0,024	- 0,057	- 0,026	- 0,016	- 0,010	- 0,024	- 0,024	0,014	5,50
5	- 0,020	- 0,009	- 0,012	+ 0,009	+ 0,016	- 0,027	- 0,021	- 0,006	+ 0,010	+ 0,021	- 0,004	0,017	0,72
6	- 0,045	- 0,019	- 0,017	+ 0,009	+ 0,006	- 0,037	- 0,006	- 0,006	- 0,000	- 0,019	- 0,013	0,018	2,41
7	+ 0,129	+ 0,032	+ 0,006	+ 0,014	- 0,004	+ 0,113	+ 0,009	+ 0,029	- 0,016	+ 0,006	+ 0,032	0,049	2,04
8	+ 0,010	+ 0,011	+ 0,008	+ 0,024	+ 0,026	+ 0,013	+ 0,014	+ 0,014	+ 0,020	+ 0,031	+ 0,017	0,008	6,97
9	- 0,050	- 0,049	- 0,072	- 0,056	- 0,074	+ 0,013	+ 0,014	+ 0,144	- 0,075	- 0,109	- 0,031	0,072	1,37
10	+ 0,015	+ 0,011	+ 0,003	+ 0,009	- 0,009	- 0,162	+ 0,009	- 0,001	- 0,005	- 0,024	- 0,015	0,053	0,92
11	- 0,025	- 0,014	- 0,012	+ 0,004	+ 0,001	+ 0,128	- 0,021	- 0,016	+ 0,035	+ 0,026	+ 0,011	0,046	0,73
12	+ 0,065	+ 0,041	+ 0,053	- 0,021	+ 0,046	+ 0,073	+ 0,064	+ 0,044	+ 0,025	+ 0,071	+ 0,046	0,028	5,21
13	- 0,090	- 0,069	- 0,062	- 0,041	- 0,034	- 0,082	- 0,051	- 0,056	- 0,040	- 0,014	- 0,054	0,023	7,43
14	+ 0,015	+ 0,006	- 0,009	- 0,008	- 0,014	+ 0,016	- 0,016	- 0,019	- 0,013	- 0,021	- 0,006	0,014	1,46
15	+ 0,035	+ 0,011	+ 0,008	+ 0,029	+ 0,001	+ 0,018	+ 0,009	- 0,001	- 0,000	+ 0,001	+ 0,011	0,013	2,77
16	+ 0,025	+ 0,016	+ 0,013	+ 0,024	+ 0,016	+ 0,033	+ 0,024	+ 0,009	+ 0,020	+ 0,031	+ 0,021	0,008	8,47
17	- 0,005	+ 0,011	+ 0,028	- 0,006	- 0,024	+ 0,008	+ 0,009	- 0,011	- 0,020	- 0,069	- 0,008	0,027	0,94
18	+ 0,405	+ 0,001	+ 0,048	+ 0,019	+ 0,021	+ 0,028	- 0,001	+ 0,024	+ 0,020	+ 0,036	+ 0,060	0,122	1,56
19	- 0,050	+ 0,036	+ 0,028	- 0,051	+ 0,016	- 0,117	- 0,046	+ 0,014	- 0,325	- 0,334	- 0,083	0,138	1,90
20	- 0,070	- 0,069	- 0,052	- 0,041	- 0,044	- 0,062	- 0,051	- 0,056	- 0,040	- 0,029	- 0,051	0,013	12,24
21	+ 0,060	+ 0,041	+ 0,038	+ 0,034	+ 0,026	+ 0,058	+ 0,034	+ 0,034	+ 0,030	+ 0,031	+ 0,039	0,012	10,50
d	+ 0,002	+ 0,002	+ 0,001	+ 0,001	- 0,001	+ 0,002	+ 0,002	+ 0,001	- 0,001	- 0,001	+ 0,000	0,059	
Sd	0,052	0,034	0,034	0,028	0,030	0,071	0,031	0,026	0,030	0,042	0,040		

d = mean of differences

Sd = standard deviation of differences

t = Student test - comparison to 0

Upper limits : $\bar{d} = \pm 0,02 \text{ g / 100 g}$ Sd = 0,03 g / 100g**ISO 9622 | IDF 141 : Precision of the method :**

Sr = 0.014 g / 100 g

SR = 0,04 g / 100 g

Table VI : Zscore of the different laboratories for each sample.
ZS calculated on the PT standard deviation

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	+0,10	+0,17	-0,06	+0,33	-0,13	+0,05	+0,60	-0,05	+0,16	-0,09
2	+0,87	+1,50	+0,83	+0,50	-0,46	+0,68	+1,40	+0,51	-0,34	-0,68
3	+0,49	+0,91	+0,83	+1,57	+1,73	+0,47	+1,24	+1,08	+1,84	+1,33
4	-0,66	-0,71	-0,65	-0,21	-0,80	-0,80	-0,84	-0,62	-0,34	-0,56
5	-0,38	-0,27	-0,35	+0,33	+0,55	-0,38	-0,68	-0,24	+0,33	+0,50
6	-0,86	-0,56	-0,50	+0,33	+0,21	-0,52	-0,20	-0,24	-0,00	-0,45
7	+2,48	+0,95	+0,16	+0,48	-0,13	+1,60	+0,28	+1,08	-0,54	+0,15
8	+0,20	+0,32	+0,24	+0,86	+0,89	+0,19	+0,44	+0,51	+0,67	+0,74
9	-0,95	-1,44	-2,13	-1,98	-2,49	+0,19	+0,44	+5,42	-2,51	-2,57
10	+0,29	+0,32	+0,09	+0,33	-0,30	-2,29	+0,28	-0,05	-0,17	-0,56
11	-0,47	-0,41	-0,35	+0,15	+0,04	+1,82	-0,68	-0,62	+1,17	+0,62
12	+1,25	+1,20	+1,57	-0,74	+1,56	+1,04	+2,05	+1,65	+0,83	+1,68
13	-1,72	-2,03	-1,83	-1,45	-1,14	-1,16	-1,64	-2,13	-1,34	-0,33
14	+0,29	+0,17	-0,28	-0,30	-0,46	+0,22	-0,52	-0,72	-0,42	-0,50
15	+0,68	+0,32	+0,24	+1,04	+0,04	+0,26	+0,28	-0,05	-0,00	+0,03
16	+0,49	+0,47	+0,38	+0,86	+0,55	+0,47	+0,76	+0,32	+0,67	+0,74
17	-0,09	+0,32	+0,83	-0,21	-0,80	+0,12	+0,28	-0,43	-0,67	-1,63
18	+7,77	+0,03	+1,42	+0,68	+0,72	+0,40	-0,04	+0,89	+0,67	+0,86
19	-0,95	+1,06	+0,83	-1,81	+0,55	-1,65	-1,48	+0,51	-10,88	-7,89
20	-1,34	-2,03	-1,54	-1,45	-1,48	-0,87	-1,64	-2,13	-1,34	-0,68
21	+1,16	+1,20	+1,12	+1,21	+0,89	+0,83	+1,08	+1,27	+1,00	+0,74

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 2 :

Zscore of the different laboratories for each sample. ZS calculated on the PT standard deviation




Table VII : Zscore of the different laboratories for each sample.
ZS calculated on the standard deviation of reproducibility of the method

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	+0,13	+0,15	-0,05	+0,23	-0,09	+0,08	+0,47	-0,04	+0,12	-0,10
2	+1,13	+1,27	+0,70	+0,35	-0,34	+1,21	+1,09	+0,34	-0,25	-0,72
3	+0,63	+0,77	+0,70	+1,10	+1,28	+0,83	+0,97	+0,71	+1,37	+1,40
4	-0,87	-0,60	-0,55	-0,15	-0,59	-1,42	-0,66	-0,41	-0,25	-0,60
5	-0,49	-0,23	-0,30	+0,23	+0,41	-0,67	-0,53	-0,16	+0,25	+0,53
6	-1,12	-0,48	-0,42	+0,23	+0,16	-0,92	-0,16	-0,16	-0,00	-0,47
7	+3,23	+0,81	+0,14	+0,34	-0,09	+2,83	+0,22	+0,71	-0,40	+0,15
8	+0,26	+0,27	+0,20	+0,60	+0,66	+0,33	+0,34	+0,34	+0,50	+0,78
9	-1,24	-1,23	-1,80	-1,40	-1,84	+0,33	+0,34	+3,59	-1,88	-2,72
10	+0,38	+0,27	+0,08	+0,23	-0,22	-4,04	+0,22	-0,04	-0,13	-0,60
11	-0,62	-0,35	-0,30	+0,10	+0,03	+3,21	-0,53	-0,41	+0,87	+0,65
12	+1,63	+1,02	+1,33	-0,52	+1,16	+1,83	+1,59	+1,09	+0,62	+1,78
13	-2,24	-1,73	-1,55	-1,02	-0,84	-2,04	-1,28	-1,41	-1,00	-0,35
14	+0,38	+0,15	-0,24	-0,21	-0,34	+0,40	-0,41	-0,47	-0,31	-0,53
15	+0,88	+0,27	+0,20	+0,73	+0,03	+0,46	+0,22	-0,04	-0,00	+0,03
16	+0,63	+0,40	+0,33	+0,60	+0,41	+0,83	+0,59	+0,21	+0,50	+0,78
17	-0,12	+0,27	+0,70	-0,15	-0,59	+0,21	+0,22	-0,29	-0,50	-1,72
18	+10,13	+0,02	+1,20	+0,48	+0,53	+0,71	-0,03	+0,59	+0,50	+0,90
19	-1,24	+0,90	+0,70	-1,27	+0,41	-2,92	-1,16	+0,34	-8,13	-8,35
20	-1,74	-1,73	-1,30	-1,02	-1,09	-1,54	-1,28	-1,41	-1,00	-0,72
21	+1,51	+1,02	+0,95	+0,85	+0,66	+1,46	+0,84	+0,84	+0,75	+0,78


This table will allows to compare your ZSCORE from one PT to an other because the standard deviation has always the value of SR of the method SR=0,040

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 3 :

Zscore of the different laboratories for each sample. ZS calculated on the standard deviation of reproducibility of the method






Figure 1 : ACCURACY - Evaluation of the individual performances (to see table I).

LIST OF ICAR PARTICIPANTS

RAW MILK

FAT ROUTINE METHOD

MARCH 2017

AIA-LAB. STANDARD LATTE	MACCARESE (ROMA)	ITALY
CATTLE INFORMATION SERVICE (CIS)	TELFORD	UNITED KINGDOM
COOPRINSEM	OSORNO	CHILE
CROATIAN AGRICULTURAL AGENCY MILK QUALITY CONTROL	KRIZEVCI	CROATIA
EASTERN LAB SERVICES	MEDINA	USA
EUROFINS STEINS LABORATORY A/B	VEJEN	DENMARK
EUROFINS STEINS LABORATORY A/B	JÖNGÖPING	SWEDEN
FED.LATTERIE SOCIALI DI BOLZANO	BOLZANO	ITALY
LAB. POLJOPRIVREDNI	NOVI SAD	SERBIA
LOM KCHZ LABORATORIUM REFERENCYJNE z/s w	PARZNIEWIE	POLAND
MERIEUX NUTRISCIENCE	CAPE TOWN	SOUTH AFRICA
MERIEUX NUTRISCIENCE	MIDRAND	SOUTH AFRICA
MERIEUX NUTRISCIENCE J BAY	CAPE TOWN	SOUTH AFRICA
MILCHPRUFRING BADEN-WURTTENBERG E.V.	KIRCHEIM / TECK	GERMANY
SHANGAI DAIRY BREEDING CENTER Co.Ltd	SHANGHAI	CHINA
TAIWAN LIVESTOCK ESEARCH INT.	TAIWAN	TAIWAN
VETLAB AGRICULTURAL SHOWGROUNDS	LUSAKA	ZAMBIA



ICAR
PROFICIENCY TESTING SCHEME

March 2017

Raw Milk

Determination of CRUDE PROTEIN CONTENT
Routine method

Sending date of statistical treatment : 21th April 2017

Frame of activity :	ICAR Milk Analyses Sub Committee (MA SC)		
Contact :	Gavin Scott	gavin@milktest.co.nz	
ICAR Staff	Silvia Orlandini	pt@icar.org	silvia@icar.org

Table I : Ranking of the laboratories

Units : g / 100 g

Nb	%	N°	d	Sd	D	Method
1	5	20	+ 0,004	0,008	0,009	IR
2	10	3	- 0,006	0,009	0,011	IR
3	14	9	- 0,010	0,005	0,012	IR
4	19	13	+ 0,014	0,010	0,017	IR
5	24	21	- 0,007	0,019	0,020	IR
6	29	19	- 0,021	0,014	0,025	IR
7	33	16	+ 0,026	0,009	0,028	IR
8	38	15	- 0,023	0,022	0,032	IR
9	43	5	+ 0,032	0,008	0,032	IR
10	48	2	- 0,010	0,037	0,038	Other*
11	52	8	+ 0,038	0,008	0,038	IR
12	57	14	+ 0,041	0,010	0,042	IR
13	62	17	- 0,043	0,017	0,047	IR
14	67	10	- 0,046	0,012	0,048	IR
15	71	4	- 0,037	0,033	0,050	IR
16	76	6	+ 0,055	0,009	0,055	IR
17	81	7	+ 0,056	0,010	0,057	IR
18	86	18	+ 0,056	0,011	0,057	IR
19	90	11	+ 0,076	0,016	0,078	IR
20	95	12	- 0,080	0,011	0,081	IR
21	100	1	- 0,136	0,017	0,137	IR

The table should be studied in parallel with figure 1 where the laboratories are located according to an acceptability area (or target) the limits of which are :

$$\pm 0,025 \text{ g / 100 g for } \bar{d} \text{ and } 0,020 \text{ g / 100 g for } S_d$$

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 20 sets of results send by 20 laboratories using routine method ISO 9622|IDF 141, after outlier discarding using Grubbs test at 5 % risk level

Other* : ASU L 01.00-78 IR-mesurement

(NC : OUT of RANKING because of insufficient data number)

(Nb : laboratory rank; % : relative rank)

(N° : laboratory identification number)

(d et Sd : mean and standard deviation of the differences (laboratory -reference))

(D : Euclidian distance to YX-axis origin = SQUARE ROOT.(d² + Sd²))

Note : Limits are only indicative and so far do not constitute standard values; they indicate what is normally reachable by labs for their self evaluation.

Repeatability standard deviation of this ICAR proficiency test (after Cochran elimination at 5 %)

Sr_{PT} 0,007

Reproducibility standard deviation of this ICAR proficiency test (after Cochran and Grubbs elimination at 5 %)

SR_{PT} 0,052

Table II : REPEATABILITY - Absolute difference between replicates in g / 100 g

Sample Lab code	1	2	3	4	5	6	7	8	9	10	Sr	NL
1	0,020	0,010	0,010	0,010	0,010	0,000	0,010	0,010	0,030 *	0,010	0,010	20
2	0,010	0,010	0,010	0,020	0,010	0,010	0,010	0,010	0,010	0,020	0,009	20
3	0,010	0,000	0,000	0,010	0,010	0,010	0,010	0,000	0,000	0,010	0,005	20
4	0,020	0,020	0,010	0,000	0,010	0,010	0,000	0,010	0,000	0,000	0,008	20
5	0,000	0,010	0,000	0,010	0,010	0,010	0,010	0,000	0,000	0,020	0,007	20
6	0,010	0,010	0,010	0,010	0,000	0,020	0,010	0,000	0,010	0,010	0,007	20
7	0,000	0,010	0,005	0,015	0,020	0,020	0,020	0,020	0,010	0,020	0,011	20
8	0,000	0,010	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,002	20
9	0,010	0,000	0,010	0,010	0,010	0,000	0,000	0,000	0,010	0,020	0,007	20
10	0,010	0,000	0,000	0,010	0,000	0,000	0,000	0,000	0,000	0,000	0,003	20
11	0,000	0,000	0,010	0,000	0,000	0,010	0,000	0,000	0,000	0,000	0,003	20
12	0,000	0,000	0,010	0,010	0,000	0,000	0,020	0,000	0,000	0,010	0,006	20
13	0,000	0,000	0,000	0,010	0,010	0,000	0,010	0,000	0,010	0,000	0,004	20
14	0,000	0,005	0,005	0,005	0,010	0,005	0,005	0,005	0,005	0,005	0,004	20
15	0,010	0,010	0,010	0,000	0,020	0,010	0,010	0,010	0,020	0,010	0,009	20
16	0,010	0,010	0,010	0,010	0,010	0,000	0,010	0,010	0,010	0,000	0,006	20
17	0,010	0,000	0,000	0,010	0,010	0,010	0,000	0,010	0,000	0,000	0,005	20
18	0,000	0,000	0,010	0,020	0,000	0,000	0,000	0,000	0,000	0,010	0,005	20
19	0,010	0,010	0,010	0,000	0,010	0,010	0,000	0,000	0,000	0,010	0,005	20
20	0,010	0,010	0,010	0,000	0,010	0,000	0,010	0,000	0,000	0,000	0,005	20
21	0,010	0,000	0,010	0,000	0,020	0,000	0,000	0,020	0,020	0,020	0,009	20
Sr	0,007	0,006	0,006	0,007	0,008	0,006	0,006	0,006	0,007	0,008		420
NE	42	42	42	42	42	42	42	42	42	42		
L	0,027	0,023	0,023	0,028	0,030	0,025	0,026	0,023	0,024	0,032		

Sr : repeatability standard deviation of each laboratory limit 0,014 g /100g

NL : number of measurements per laboratory

L : Limit for difference between duplicates according Cochran test at 5% level.

SE : repeatability standard deviation per sample

NE : number of measurements per sample

*: discarded data using the test of Cochran at 5 %

** : missing data

r : limit of repeatability, absolute difference between two replicates=0,040 according ISO 9622 |IDF 141

Table III : Means of the replicates in g / 100 g

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	3,600	3,425	3,135	2,775	2,585	3,500	3,325 *	2,955	2,675	3,115
2	3,695	3,535	3,245	2,900	2,725	3,605	3,445	3,075	2,825	3,300
3	3,765	3,560	3,260	2,885	2,675	3,665	3,475	3,070	2,780	3,255
4	3,760	3,570	3,235	2,830	2,605	3,655	3,460	3,025	2,720	3,220
5	3,780	3,605	3,300	2,935	2,735	3,685	3,495	3,110	2,830	3,290
6	3,805	3,615	3,325	2,955	2,760	3,700	3,525	3,140	2,865	3,305
7	3,810	3,615	3,328	2,963	2,760	3,700	3,520	3,140	2,865	3,310
8	3,810	3,615	3,310	2,930	2,720	3,700	3,510	3,110	2,830	3,290
9	3,745	3,560	3,255	2,885	2,685	3,640	3,460	3,070	2,795	3,250
10	3,705	3,520	3,220	2,855	2,660	3,580	3,430	3,040	2,760	3,220
11	3,860	3,660	3,345	2,960	2,750	3,745	3,560	3,150	2,850	3,330
12	3,660	3,480	3,195	2,825	2,620	3,570	3,390	3,010	2,730	3,165
13	3,780	3,590	3,290	2,895	2,695	3,680	3,495	3,090	2,795	3,280
14	3,795	3,613	3,308	2,938	2,745	3,693	3,498	3,118	2,848	3,303
15	3,705	3,525	3,245	2,880	2,690	3,615	3,435	3,055	2,810	3,255
16	3,795	3,605	3,295	2,915	2,705	3,690	3,495	3,105	2,815	3,290
17	3,735	3,540	3,230	2,845	2,635	3,635	3,440	3,035	2,740	3,180
18	3,810	3,640	3,335	2,940	2,730	3,730	3,540	3,130	2,850	3,305
19	3,715	3,535	3,245	2,890	2,675	3,625	3,450	3,070	2,800	3,235
20	3,765	3,575	3,275	2,890	2,685	3,670	3,485	3,080	2,790	3,270
21	3,735	3,550	3,265	2,910	2,710	3,620	3,450	3,080	2,810	3,250
M	3,754	3,568	3,269	2,895	2,693	3,653	3,478	3,079	2,799	3,258
REF.	3,757	3,570	3,270	2,895	2,691	3,655	3,480	3,079	2,798	3,256
SD	0,060	0,055	0,051	0,049	0,050	0,058	0,042	0,048	0,051	0,054

M = mean per sample

REF. = reference values

SD = standard deviation per sample

*: discarded data using the test of Grubbs 5 %

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 20 laboratories using the Routine method ISO 9622 | IDF 141, after outliers discarding using Grubbs test at 5 % risk level.

Table IV : Outlier identification

Sample	1	2	3	4	5	6	7	8	9	10
Outliers Cochran									1	
Outlier Grubbs							1			
sr	0,007	0,006	0,006	0,007	0,008	0,006	0,006	0,006	0,006	0,008
SR	0,060	0,056	0,051	0,049	0,050	0,058	0,042	0,048	0,043	0,054

Table V : ACCURACY - differences (laboratory - reference) in g / 100 g

Sample Lab code	1	2	3	4	5	6	7	8	9	10	d	Sd _{lab}	t
1	- 0,157	- 0,145	- 0,135	- 0,120	- 0,106	- 0,155	- 0,155	- 0,124	- 0,123	- 0,141	- 0,136	0,017	24,87
2	- 0,062	- 0,035	- 0,025	+ 0,005	+ 0,034	- 0,050	- 0,035	- 0,004	+ 0,027	+ 0,044	- 0,010	0,037	0,86
3	+ 0,008	- 0,010	- 0,010	- 0,010	- 0,016	+ 0,010	- 0,005	- 0,009	- 0,018	- 0,001	- 0,006	0,009	2,02
4	+ 0,003	+ 0,000	- 0,035	- 0,065	- 0,086	+ 0,000	- 0,020	- 0,054	- 0,078	- 0,036	- 0,037	0,033	3,53
5	+ 0,023	+ 0,035	+ 0,030	+ 0,040	+ 0,044	+ 0,030	+ 0,015	+ 0,031	+ 0,032	+ 0,034	+ 0,032	0,008	12,47
6	+ 0,048	+ 0,045	+ 0,055	+ 0,060	+ 0,069	+ 0,045	+ 0,045	+ 0,061	+ 0,067	+ 0,049	+ 0,055	0,009	18,75
7	+ 0,053	+ 0,045	+ 0,058	+ 0,067	+ 0,069	+ 0,045	+ 0,040	+ 0,061	+ 0,067	+ 0,054	+ 0,056	0,010	17,36
8	+ 0,053	+ 0,045	+ 0,040	+ 0,035	+ 0,029	+ 0,045	+ 0,030	+ 0,031	+ 0,032	+ 0,034	+ 0,038	0,008	14,65
9	- 0,012	- 0,010	- 0,015	- 0,010	- 0,006	- 0,015	- 0,020	- 0,009	- 0,003	- 0,006	- 0,010	0,005	6,69
10	- 0,052	- 0,050	- 0,050	- 0,040	- 0,031	- 0,075	- 0,050	- 0,039	- 0,038	- 0,036	- 0,046	0,012	11,75
11	+ 0,103	+ 0,090	+ 0,075	+ 0,065	+ 0,059	+ 0,090	+ 0,080	+ 0,071	+ 0,052	+ 0,074	+ 0,076	0,016	15,44
12	- 0,097	- 0,090	- 0,075	- 0,070	- 0,071	- 0,085	- 0,090	- 0,069	- 0,068	- 0,091	- 0,080	0,011	23,25
13	+ 0,023	+ 0,020	+ 0,020	- 0,000	+ 0,004	+ 0,025	+ 0,015	+ 0,011	- 0,003	+ 0,024	+ 0,014	0,010	4,23
14	+ 0,038	+ 0,043	+ 0,038	+ 0,042	+ 0,054	+ 0,038	+ 0,018	+ 0,038	+ 0,050	+ 0,047	+ 0,041	0,010	13,25
15	- 0,052	- 0,045	- 0,025	- 0,015	- 0,001	- 0,040	- 0,045	- 0,024	+ 0,012	- 0,001	- 0,023	0,022	3,39
16	+ 0,038	+ 0,035	+ 0,025	+ 0,020	+ 0,014	+ 0,035	+ 0,015	+ 0,026	+ 0,017	+ 0,034	+ 0,026	0,009	8,95
17	- 0,022	- 0,030	- 0,040	- 0,050	- 0,056	- 0,020	- 0,040	- 0,044	- 0,058	- 0,076	- 0,043	0,017	7,93
18	+ 0,053	+ 0,070	+ 0,065	+ 0,045	+ 0,039	+ 0,075	+ 0,060	+ 0,051	+ 0,052	+ 0,049	+ 0,056	0,011	15,43
19	- 0,042	- 0,035	- 0,025	- 0,005	- 0,016	- 0,030	- 0,030	- 0,009	+ 0,002	- 0,021	- 0,021	0,014	4,77
20	+ 0,008	+ 0,005	+ 0,005	- 0,005	- 0,006	+ 0,015	+ 0,005	+ 0,001	- 0,008	+ 0,014	+ 0,004	0,008	1,38
21	- 0,022	- 0,020	- 0,005	+ 0,015	+ 0,019	- 0,035	- 0,030	+ 0,001	+ 0,012	- 0,006	- 0,007	0,019	1,16
d	- 0,003	- 0,002	- 0,001	+ 0,000	+ 0,002	- 0,002	- 0,002	- 0,000	+ 0,001	+ 0,002	- 0,001	0,052	
Sd	0,060	0,055	0,051	0,049	0,050	0,058	0,042	0,048	0,051	0,054	0,052		

d = mean of differences

Sd = standard deviation of differences

t = Student test - comparison to 0

Upper limits : $\bar{d} = \pm 0,025 \text{ g / 100 g}$

Sd = 0,020 g / 100 g

ISO 9622|IDF141 : Precision of the method :

Sr = 0.014 g / 100 g

SR = 0,04 g / 100 g

Table VI : Zscore of the different laboratories for each sample.
ZS calculated on the PT standard deviation

Sample Lab Code	1	2	3	4	5	6	7	8	9	10
1	-2,62	-2,62	-2,63	-2,45	-2,12	-2,66	-3,68	-2,57	-2,43	-2,60
2	-1,03	-0,63	-0,48	+0,10	+0,67	-0,86	-0,82	-0,09	+0,54	+0,82
3	+0,14	-0,18	-0,19	-0,20	-0,32	+0,17	-0,11	-0,19	-0,35	-0,02
4	+0,05	+0,00	-0,68	-1,33	-1,72	+0,00	-0,47	-1,12	-1,54	-0,66
5	+0,39	+0,63	+0,59	+0,82	+0,87	+0,52	+0,37	+0,64	+0,63	+0,63
6	+0,81	+0,82	+1,08	+1,23	+1,37	+0,77	+1,08	+1,26	+1,33	+0,91
7	+0,89	+0,82	+1,13	+1,38	+1,37	+0,77	+0,96	+1,26	+1,33	+1,00
8	+0,89	+0,82	+0,79	+0,72	+0,57	+0,77	+0,72	+0,64	+0,63	+0,63
9	-0,20	-0,18	-0,29	-0,20	-0,12	-0,26	-0,47	-0,19	-0,06	-0,11
10	-0,87	-0,90	-0,97	-0,82	-0,62	-1,29	-1,18	-0,81	-0,75	-0,66
11	+1,73	+1,63	+1,47	+1,33	+1,17	+1,55	+1,91	+1,47	+1,03	+1,37
12	-1,62	-1,62	-1,46	-1,43	-1,42	-1,46	-2,13	-1,43	-1,34	-1,68
13	+0,39	+0,36	+0,40	-0,00	+0,07	+0,43	+0,37	+0,23	-0,06	+0,45
14	+0,64	+0,77	+0,74	+0,87	+1,07	+0,65	+0,43	+0,79	+0,98	+0,86
15	-0,87	-0,81	-0,48	-0,31	-0,02	-0,68	-1,06	-0,50	+0,24	-0,02
16	+0,64	+0,63	+0,49	+0,41	+0,27	+0,60	+0,37	+0,54	+0,34	+0,63
17	-0,36	-0,54	-0,78	-1,02	-1,12	-0,34	-0,94	-0,91	-1,14	-1,40
18	+0,89	+1,27	+1,27	+0,92	+0,77	+1,29	+1,44	+1,05	+1,03	+0,91
19	-0,70	-0,63	-0,48	-0,10	-0,32	-0,51	-0,70	-0,19	+0,04	-0,39
20	+0,14	+0,09	+0,10	-0,10	-0,12	+0,26	+0,13	+0,02	-0,16	+0,26
21	-0,36	-0,36	-0,09	+0,31	+0,37	-0,60	-0,70	+0,02	+0,24	-0,11

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 2 :

Zscore of the different laboratories for each sample. ZS calculated on the PT standard deviation




Table VII : Zscore of the different laboratories for each sample.
ZS calculated on the standard deviation of reproducibility of the method

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	-3,92	-3,62	-3,37	-3,00	-2,66	-3,87	-3,87	-3,10	-3,07	-3,52
2	-1,54	-0,87	-0,62	+0,12	+0,84	-1,25	-0,87	-0,10	+0,68	+1,10
3	+0,21	-0,25	-0,24	-0,25	-0,41	+0,25	-0,12	-0,23	-0,45	-0,02
4	+0,08	+0,00	-0,87	-1,63	-2,16	+0,00	-0,49	-1,35	-1,95	-0,90
5	+0,58	+0,88	+0,76	+1,00	+1,09	+0,75	+0,38	+0,77	+0,80	+0,85
6	+1,21	+1,13	+1,38	+1,50	+1,72	+1,13	+1,13	+1,52	+1,68	+1,23
7	+1,33	+1,13	+1,44	+1,69	+1,72	+1,13	+1,01	+1,52	+1,68	+1,35
8	+1,33	+1,13	+1,01	+0,87	+0,72	+1,13	+0,76	+0,77	+0,80	+0,85
9	-0,29	-0,25	-0,37	-0,25	-0,16	-0,37	-0,49	-0,23	-0,07	-0,15
10	-1,29	-1,25	-1,24	-1,00	-0,78	-1,87	-1,24	-0,98	-0,95	-0,90
11	+2,58	+2,25	+1,88	+1,62	+1,47	+2,25	+2,01	+1,77	+1,30	+1,85
12	-2,42	-2,25	-1,87	-1,75	-1,78	-2,12	-2,24	-1,73	-1,70	-2,27
13	+0,58	+0,50	+0,51	-0,00	+0,09	+0,63	+0,38	+0,27	-0,07	+0,60
14	+0,96	+1,07	+0,94	+1,06	+1,34	+0,94	+0,45	+0,96	+1,24	+1,17
15	-1,29	-1,12	-0,62	-0,38	-0,03	-1,00	-1,12	-0,60	+0,30	-0,02
16	+0,96	+0,88	+0,63	+0,50	+0,34	+0,88	+0,38	+0,65	+0,43	+0,85
17	-0,54	-0,75	-0,99	-1,25	-1,41	-0,50	-0,99	-1,10	-1,45	-1,90
18	+1,33	+1,75	+1,63	+1,13	+0,97	+1,88	+1,51	+1,27	+1,30	+1,23
19	-1,04	-0,87	-0,62	-0,13	-0,41	-0,75	-0,74	-0,23	+0,05	-0,52
20	+0,21	+0,13	+0,13	-0,13	-0,16	+0,38	+0,13	+0,02	-0,20	+0,35
21	-0,54	-0,50	-0,12	+0,37	+0,47	-0,87	-0,74	+0,02	+0,30	-0,15


This table will allows to compare your ZSCORE from one PT to an other because the standard deviation has always the value of SR of the method SR=0,040

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 3 :

Zscore of the different laboratories for each sample. ZS calculated on the standard deviation of reproducibility of the method






Figure 1 : ACCURACY - Evaluation of the individual performances (to see table I).

LIST OF ICAR PARTICIPANTS
RAW MILK
PROTEIN ROUTINE METHOD
MARCH 2017

AIA-LAB. STANDARD LATTE	MACCARESE (ROMA)	ITALY
CATTLE INFORMATION SERVICE (CIS)	TELFORD	UNITED KINGDOM
COOPRINSEM	OSORNO	CHILE
CROATIAN AGRICULTURAL AGENCY MILK QUALITY CONTROL	KRIZEVCI	CROATIA
EASTERN LAB SERVICES	MEDINA	USA
EUROFINS STEINS LABORATORY A/B	VEJEN	DENMARK
EUROFINS STEINS LABORATORY A/B	JÖNGÖPING	SWEDEN
FED.LATTERIE SOCIALI DI BOLZANO	BOLZANO	ITALY
LAB. POLJOPRIVREDNI	NOVI SAD	SERBIA
LOM KCHZ LABORATORIUM REFERENCYJNE z/s w	PARZNIEWIE	POLAND
MERIEUX NUTRISCIENCE	CAPE TOWN	SOUTH AFRICA
MERIEUX NUTRISCIENCE	MIDRAND	SOUTH AFRICA
MERIEUX NUTRISCIENCE J BAY	CAPE TOWN	SOUTH AFRICA
MILCHPRUFRING BADEN-WURTTENBERG E.V.	KIRCHHEIM / TECK	GERMANY
SHANGAI DAIRY BREEDING CENTER Co.Ltd	SHANGHAI	CHINA
TAIWAN LIVESTOCK RESEARCH INT.	TAIWAN	TAIWAN
VETLAB AGRICULTURAL SHOWGROUNDS	LUSAKA	ZAMBIA



**ICAR
PROFICIENCY TESTING SCHEME**

March 2017

Raw Milk

**Determination of LACTOSE CONTENT
Routine method**

Sending date of statistical treatment : 21th April 2017

Frame of activity :	ICAR Milk Analyses Sub Committee (MA SC)		
Contact :	Gavin Scott	gavin@milktest.co.nz	
ICAR Staff	Silvia Orlandini	pt@icar.org	silvia@icar.org

Table I : **Ranking of the laboratories** Units : g / 100 g

Nb	%	N°	d	Sd	D	Method
1	6	14	- 0,019	0,010	0,021	IR
2	11	4	- 0,007	0,021	0,022	IR
3	17	18	- 0,029	0,010	0,030	IR
4	22	10	- 0,007	0,030	0,031	IR
5	28	13	- 0,034	0,026	0,043	IR
6	33	17	+ 0,038	0,029	0,048	IR
7	39	3	+ 0,008	0,055	0,056	IR
8	44	11	- 0,066	0,009	0,067	IR
9	50	5	- 0,068	0,010	0,068	IR
10	56	15	+ 0,080	0,006	0,080	IR
11	61	9	- 0,081	0,009	0,082	IR
12	67	8	+ 0,094	0,021	0,096	IR
13	72	2	- 0,098	0,011	0,099	Other*
14	78	1	- 0,104	0,023	0,107	IR
15	83	7	+ 0,108	0,007	0,108	IR
16	89	16	+ 0,115	0,008	0,115	IR
17	94	12	+ 0,120	0,010	0,120	IR
18	100	6	- 0,149	0,006	0,149	IR

The table should be studied in parallel with figure 1 where the laboratories are located according to an acceptability area (or target) the limits of which are :

$$\pm 0.100 \text{ g / 100 g for } \bar{d} \text{ and } 0.100 \text{ g / 100g for } Sd$$

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 17 sets of results send by 17 laboratories using routine method ISO 9622|IDF 141, after outlier discarding using Grubbs test at 5 % risk level

Other* : ASU L 01.00-78 IR-mesurement

(NC : OUT of RANKING because of insufficient data number)

(Nb : laboratory rank; % : relative rank)

(N° : laboratory identification number)

(d et Sd : mean and standard deviation of the differences (laboratory -reference))

(D : Euclidian distance to YX-axis origin = SQUARE ROOT.(d² + Sd²))

Note : Limits are only indicative and so far do not constitute standard values; they indicate what is normally reachable by labs for their self evaluation.

Repeatability standard deviation of this ICAR proficiency test (after Cochran elimination at 5 %)

Sr_{PT} 0,007

Reproducibility standard deviation of this ICAR proficiency test (after Cochran and Grubbs elimination at 5 %)

SR_{PT} 0,085

Table II : REPEATABILITY - Absolute difference between replicates in g / 100 g

Sample Lab Code	1	2	3	4	5	6	7	8	9	10	Sr	NL
1	0,010	0,020	0,010	0,010	0,010	0,020	0,000	0,010	0,010	0,010	0,009	20
2	0,020	0,010	0,010	0,010	0,020	0,020	0,010	0,010	0,020	0,020	0,011	20
3	0,010	0,010	0,010	0,010	0,000	0,010	0,010	0,010	0,000	0,010	0,006	20
4	0,010	0,000	0,010	0,000	0,000	0,000	0,000	0,000	0,020	0,000	0,005	20
5	0,001	0,000	0,010	0,010	0,000	0,011	0,000	0,011	0,011	0,011	0,006	20
6	0,010	0,010	0,010	0,010	0,000	0,010	0,000	0,010	0,010	0,000	0,006	20
7	0,010	0,010	0,010	0,020	0,020	0,020	0,000	0,010	0,000	0,010	0,009	20
8	0,020	0,030	0,010	0,020	0,000	0,000	0,005	0,000	0,005	0,003	0,010	20
9	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,010	0,000	0,000	0,002	20
10	0,000	0,000	0,000	0,000	0,000	0,010	0,010	0,000	0,010	0,010	0,004	20
11	0,010	0,010	0,000	0,000	0,000	0,010	0,020	0,010	0,000	0,010	0,007	20
12	0,010	0,000	0,010	0,000	0,000	0,000	0,010	0,010	0,000	0,010	0,005	20
13	0,010	0,020	0,010	0,010	0,000	0,000	0,010	0,010	0,010	0,010	0,007	20
14	0,010	0,010	0,010	0,000	0,000	0,000	0,010	0,010	0,000	0,020	0,007	20
15	0,010	0,010	0,010	0,010	0,010	0,010	0,010	0,010	0,000	0,010	0,007	20
16	0,010	0,010	0,010	0,010	0,000	0,010	0,010	0,000	0,010	0,000	0,006	20
17	0,000	0,000	0,010	0,010	0,010	0,010	0,020	0,010	0,020	0,010	0,008	20
18	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	0,000	20
Sr	0,007	0,008	0,006	0,007	0,006	0,007	0,007	0,006	0,007	0,007		360
NE	36	36	36	36	36	36	36	36	36	36		
L	0,028	0,032	0,024	0,027	0,021	0,029	0,026	0,023	0,028	0,028		

Sr : repeatability standard deviation of each laboratory limit 0,014 g/100g

NL : number of measurements per laboratory

L : Limit for difference between duplicates according Cochran test at 5% level.

SE : repeatability standard deviation per sample

NE : number of measurements per sample

*: discarded data using the test of Cochran at 5 %

** : missing data

r : limit of repeatability, absolute difference between two replicates=0,040 according ISO 9622|IDF 141

Table III : Means of the replicates in g / 100 g

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	4,965	5,010	4,865	4,895	4,785	4,810	4,740	4,755	4,625	4,715
2	5,000	5,055	4,865	4,935	4,780	4,800	4,725	4,745	4,610	4,710
3	5,035	5,085	4,935	4,995	4,890	4,895	4,865	4,875	4,840	4,875
4	5,105	5,170	4,965	5,030	4,860	4,890	4,810	4,830	4,680	4,800
5	5,029	5,092	4,897	4,960	4,810	4,824	4,734	4,771	4,687	4,729
6	4,945	5,015	4,815	4,875	4,720	4,745	4,670	4,685	4,595	4,650
7	5,205	5,275	5,065	5,140	4,980	5,000	4,920	4,945	4,850	4,905
8	5,210	5,275	5,055	5,130	4,960	4,990	4,908	4,940	4,783	4,892
9	5,020	5,080	4,880	4,950	4,790	4,820	4,730	4,755	4,660	4,710
10	5,120	5,210	4,960	5,030	4,860	4,885	4,785	4,810	4,705	4,775
11	5,035	5,105	4,890	4,960	4,810	4,835	4,740	4,775	4,660	4,735
12	5,215	5,290	5,075	5,150	5,000	5,010	4,925	4,955	4,870	4,915
13	5,035	5,080	4,905	4,975	4,850	4,860	4,815	4,835	4,695	4,815
14	5,065	5,155	4,945	5,010	4,860	4,880	4,785	4,815	4,720	4,780
15	5,165	5,245	5,035	5,105	4,945	4,975	4,895	4,925	4,830	4,885
16	5,215	5,285	5,075	5,145	4,980	5,005	4,925	4,950	4,855	4,920
17	5,100	5,160	4,985	5,035	4,915	4,925	4,900	4,915	4,770	4,885
18	5,070	5,140	4,940	5,000	4,840	4,870	4,780	4,810	4,710	4,760
M	5,085	5,152	4,953	5,018	4,869	4,890	4,814	4,838	4,730	4,803
REF.	5,090	5,157	4,958	5,023	4,874	4,895	4,819	4,844	4,737	4,809
SD	0,087	0,093	0,081	0,086	0,081	0,080	0,083	0,084	0,090	0,086

M = mean per sample

REF. = reference values

SD = standard deviation per sample

*: discarded data using the test of Grubbs 5 %

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 17 laboratories using the Routine method ISO 9622|IDF 141 , after outliers discarding using Grubbs test at 5 % risk level.

Table IV : Outlier identification

Sample	1	2	3	4	5	6	7	8	9	10
Outliers										
Cochran										
Outlier										
Grubbs										
sr	0,007	0,008	0,006	0,007	0,006	0,007	0,007	0,006	0,007	0,007
SR	0,087	0,094	0,081	0,086	0,081	0,080	0,083	0,084	0,090	0,086

Table V : ACCURACY - differences (laboratory - reference) in g / 100 g

Sample Lab Code	1	2	3	4	5	6	7	8	9	10	d	Sd _{lab}	t
1	- 0,125	- 0,147	- 0,093	- 0,128	- 0,089	- 0,085	- 0,079	- 0,089	- 0,112	- 0,094	- 0,104	0,023	14,56
2	- 0,090	- 0,102	- 0,093	- 0,088	- 0,094	- 0,095	- 0,094	- 0,099	- 0,127	- 0,099	- 0,098	0,011	27,95
3	- 0,055	- 0,072	- 0,023	- 0,028	+ 0,016	- 0,000	+ 0,046	+ 0,031	+ 0,103	+ 0,066	+ 0,008	0,055	0,48
4	+ 0,015	+ 0,013	+ 0,007	+ 0,007	- 0,014	- 0,005	- 0,009	- 0,014	- 0,057	- 0,009	- 0,007	0,021	1,00
5	- 0,062	- 0,065	- 0,061	- 0,063	- 0,064	- 0,072	- 0,085	- 0,073	- 0,051	- 0,080	- 0,068	0,010	21,14
6	- 0,145	- 0,142	- 0,143	- 0,148	- 0,154	- 0,150	- 0,149	- 0,159	- 0,142	- 0,159	- 0,149	0,006	75,01
7	+ 0,115	+ 0,118	+ 0,107	+ 0,117	+ 0,106	+ 0,105	+ 0,101	+ 0,101	+ 0,113	+ 0,096	+ 0,108	0,007	45,84
8	+ 0,120	+ 0,118	+ 0,097	+ 0,107	+ 0,086	+ 0,095	+ 0,088	+ 0,096	+ 0,045	+ 0,083	+ 0,094	0,021	14,00
9	- 0,070	- 0,077	- 0,078	- 0,073	- 0,084	- 0,075	- 0,089	- 0,089	- 0,077	- 0,099	- 0,081	0,009	29,06
10	+ 0,030	+ 0,053	+ 0,002	+ 0,007	- 0,014	- 0,010	- 0,034	- 0,034	- 0,032	- 0,034	- 0,007	0,030	0,70
11	- 0,055	- 0,052	- 0,068	- 0,063	- 0,064	- 0,060	- 0,079	- 0,069	- 0,077	- 0,074	- 0,066	0,009	23,24
12	+ 0,125	+ 0,133	+ 0,117	+ 0,127	+ 0,126	+ 0,115	+ 0,106	+ 0,111	+ 0,133	+ 0,106	+ 0,120	0,010	37,09
13	- 0,055	- 0,077	- 0,053	- 0,048	- 0,024	- 0,035	- 0,004	- 0,009	- 0,042	+ 0,006	- 0,034	0,026	4,11
14	- 0,025	- 0,002	- 0,013	- 0,013	- 0,014	- 0,015	- 0,034	- 0,029	- 0,017	- 0,029	- 0,019	0,010	6,18
15	+ 0,075	+ 0,088	+ 0,077	+ 0,082	+ 0,071	+ 0,080	+ 0,076	+ 0,081	+ 0,093	+ 0,076	+ 0,080	0,006	39,20
16	+ 0,125	+ 0,128	+ 0,117	+ 0,122	+ 0,106	+ 0,110	+ 0,106	+ 0,106	+ 0,118	+ 0,111	+ 0,115	0,008	44,28
17	+ 0,010	+ 0,003	+ 0,027	+ 0,012	+ 0,041	+ 0,030	+ 0,081	+ 0,071	+ 0,033	+ 0,076	+ 0,038	0,029	4,26
18	- 0,020	- 0,017	- 0,018	- 0,023	- 0,034	- 0,025	- 0,039	- 0,034	- 0,027	- 0,049	- 0,029	0,010	8,92
d	- 0,005	- 0,006	- 0,005	- 0,005	- 0,005	- 0,005	- 0,005	- 0,005	- 0,007	- 0,005	- 0,005	0,083	
Sd	0,087	0,093	0,081	0,086	0,081	0,080	0,083	0,084	0,090	0,086	0,085		

d = mean of differences

Sd = standard deviation of differences

t = Student test - comparison to 0

Upper limits : $\bar{d} = \pm 0.100 \text{ g / 100g}$ Sd = 0.100 g / 100g**ISO 9622/IDF141 : Precision of the method :**

Sr = 0.014 g / 100 g

SR = 0,04 g / 100 g

Table VI : Zscore of the different laboratories for each sample.
ZS calculated on the PT standard deviation

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	-1,44	-1,57	-1,15	-1,48	-1,10	-1,07	-0,95	-1,06	-1,25	-1,09
2	-1,04	-1,09	-1,15	-1,02	-1,17	-1,20	-1,14	-1,18	-1,42	-1,15
3	-0,63	-0,77	-0,29	-0,32	+0,20	-0,00	+0,55	+0,37	+1,15	+0,78
4	+0,17	+0,14	+0,09	+0,09	-0,17	-0,07	-0,11	-0,17	-0,64	-0,10
5	-0,71	-0,70	-0,76	-0,73	-0,79	-0,90	-1,03	-0,87	-0,57	-0,93
6	-1,67	-1,52	-1,77	-1,72	-1,91	-1,89	-1,80	-1,89	-1,59	-1,85
7	+1,32	+1,26	+1,33	+1,36	+1,32	+1,32	+1,21	+1,21	+1,26	+1,12
8	+1,38	+1,26	+1,20	+1,25	+1,07	+1,19	+1,06	+1,15	+0,50	+0,97
9	-0,81	-0,83	-0,97	-0,84	-1,04	-0,95	-1,07	-1,06	-0,86	-1,15
10	+0,34	+0,57	+0,02	+0,09	-0,17	-0,13	-0,41	-0,40	-0,36	-0,39
11	-0,63	-0,56	-0,84	-0,73	-0,79	-0,76	-0,95	-0,82	-0,86	-0,86
12	+1,43	+1,42	+1,45	+1,48	+1,57	+1,44	+1,27	+1,33	+1,48	+1,24
13	-0,63	-0,83	-0,66	-0,55	-0,30	-0,44	-0,05	-0,11	-0,47	+0,08
14	-0,29	-0,02	-0,16	-0,15	-0,17	-0,19	-0,41	-0,34	-0,19	-0,33
15	+0,86	+0,94	+0,95	+0,96	+0,88	+1,00	+0,91	+0,97	+1,04	+0,89
16	+1,43	+1,37	+1,45	+1,42	+1,32	+1,38	+1,27	+1,27	+1,31	+1,30
17	+0,11	+0,03	+0,33	+0,14	+0,51	+0,37	+0,97	+0,85	+0,37	+0,89
18	-0,23	-0,18	-0,22	-0,26	-0,42	-0,32	-0,47	-0,40	-0,30	-0,57

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 2 : Zscore of the different laboratories for each sample. ZS calculated on the PT standard deviation




Table VII : Zscore of the different laboratories for each sample.
ZS calculated on the standard deviation of reproducibility of the method

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	-3,13	-3,68	-2,33	-3,19	-2,22	-2,13	-1,98	-2,22	-2,81	-2,34
2	-2,26	-2,55	-2,33	-2,19	-2,35	-2,38	-2,36	-2,47	-3,18	-2,46
3	-1,38	-1,80	-0,58	-0,69	+0,40	-0,01	+1,14	+0,78	+2,57	+1,66
4	+0,37	+0,32	+0,17	+0,18	-0,35	-0,13	-0,23	-0,35	-1,43	-0,21
5	-1,54	-1,63	-1,53	-1,57	-1,60	-1,79	-2,13	-1,83	-1,27	-2,00
6	-3,63	-3,55	-3,58	-3,69	-3,85	-3,76	-3,73	-3,97	-3,56	-3,96
7	+2,87	+2,95	+2,67	+2,93	+2,65	+2,62	+2,52	+2,53	+2,82	+2,41
8	+2,99	+2,95	+2,42	+2,68	+2,15	+2,37	+2,21	+2,40	+1,13	+2,07
9	-1,76	-1,93	-1,95	-1,82	-2,10	-1,88	-2,23	-2,22	-1,93	-2,46
10	+0,74	+1,32	+0,05	+0,18	-0,35	-0,26	-0,86	-0,85	-0,81	-0,84
11	-1,38	-1,30	-1,70	-1,57	-1,60	-1,51	-1,98	-1,72	-1,93	-1,84
12	+3,12	+3,32	+2,92	+3,18	+3,15	+2,87	+2,64	+2,78	+3,32	+2,66
13	-1,38	-1,93	-1,33	-1,19	-0,60	-0,88	-0,11	-0,22	-1,06	+0,16
14	-0,63	-0,05	-0,33	-0,32	-0,35	-0,38	-0,86	-0,72	-0,43	-0,71
15	+1,87	+2,20	+1,92	+2,06	+1,78	+1,99	+1,89	+2,03	+2,32	+1,91
16	+3,12	+3,20	+2,92	+3,06	+2,65	+2,74	+2,64	+2,65	+2,94	+2,79
17	+0,24	+0,07	+0,67	+0,31	+1,03	+0,74	+2,02	+1,78	+0,82	+1,91
18	-0,51	-0,43	-0,45	-0,57	-0,85	-0,63	-0,98	-0,85	-0,68	-1,21


This table will allows to compare your ZSCORE from one PT to an other because the standard deviation has always the value of SR of the method SR=0,040

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 3 :

Zscore of the different laboratories for each sample. ZS calculated on the standard deviation of reproducibility of the method






Figure 1 : ACCURACY - Evaluation of the individual performances (to see table I).

LIST OF PARTICIPANT LABORATORIES
RAW MILK
LACTOSE ROUTINE METHOD
MARCH 2017

AIA-LAB. STANDARD LATTE
CATTLE INFORMATION SERVICE (CIS)
CENTRAL MILK LABORATORY ICBA
EASTERN LAB SERVICES
EUROFINS STEINS LABORATORY A/B
FED.LATTERIE SOCIALI DI BOLZANO
LAB. POLJOPRIVREDNI
LOM KCHZ LABORATORIUM REFERENCYJNE z/s w
MERIEUX NUTRISCIENCE
MERIEUX NUTRISCIENCE
MERIEUX NUTRISCIENCE J BAY
MILCHPRUFUNG BADEN-WURTTENBERG E.V.
SHANGAI DAIRY BREEDING CENTER Co.Ltd
TAIWAN LIVESTOCK RESEARCH INT.
VETLAB AGRICULTURAL SHOWGROUNDS

MACCARESE (ROMA)
TELFORD
CAESAREA
MEDINA
VEJEN
BOLZANO
NOVI SAD
PARZNIEWIE
CAPE TOWN
MIDRAND
CAPE TOWN
KIRCHEIM / TECK
SHANGHAI
TAIWAN
LUSAKA

ITALY
UNITED KINGDOM
ISRAEL
USA
DENMARK
ITALY
SERBIA
POLAND
SOUTH AFRICA
SOUTH AFRICA
SOUTH AFRICA
GERMANY
CHINA
TAIWAN
ZAMBIA



ICAR
PROFICIENCY TESTING SCHEME

March 2017

Raw Milk

Determination of UREA CONTENT
Routine method

Sending date of statistical treatment : 21th April 2017

Frame of activity :	ICAR Milk Analyses Sub Committee (MA SC)		
Contact :	Gavin Scott	gavin@milktest.co.nz	
ICAR Staff	Silvia Orlandini	pt@icar.org	silvia@icar.org

Table I : **Ranking of the laboratories** Units : mg / 100 g

Nb	%	N°	d	Sd	D	Method
1	7	11	+ 0,94	0,96	1,35	Other*
2	13	1	+ 0,24	1,40	1,43	IR
3	20	14	+ 1,15	0,93	1,48	IR
4	27	6	+ 1,60	0,68	1,74	IR
5	33	2	- 2,19	0,40	2,23	IR
6	40	12	+ 2,16	0,73	2,28	IR
7	47	10	+ 1,58	1,65	2,28	IR
8	53	15	+ 2,17	1,00	2,39	IR
9	60	13	- 1,14	2,95	3,17	IR
10	67	9	- 3,57	1,18	3,75	IR
11	73	3	+ 3,76	2,40	4,46	IR
12	80	5	- 4,47	0,93	4,57	IR
13	87	7	+ 4,60	1,28	4,77	IR
14	93	8	- 5,45	10,32	11,67	IR
15	100	4	- 1,28	12,89	12,95	IR

The table should be studied in parallel with figure 1 where the laboratories are located according to an acceptability area (or target) the limits of which are :

$\pm 2,50$ mg / 100 g for $\bar{d}$ and 1,50 mg / 100 g for Sd

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 14 sets of results send by 14 laboratories using routine method ISO 9622|IDF 141, after outlier discarding using Grubbs test at 5 % risk level

Other* : home method based on ASU L 01.00.78 (ISO 9622)

(NC : OUT of RANKING because of insufficient data number)

(Nb : laboratory rank; % : relative rank)

(N° : laboratory identification number)

(d et Sd : mean and standard deviation of the differences (laboratory -reference))

(D : Euclidian distance to YX-axis origin = SQUARE ROOT.(d² + Sd²))

Sr_{PT} 0,99

SR_{PT} 3,34

Note : Limits are only indicative and so far do not constitute standard values; they indicate what is normally reachable by labs for their self evaluation.

Repeatability standard deviation of this ICAR proficiency test (after Cochran elimination at 5 %)

Reproducibility standard deviation of this ICAR proficiency test (after Cochran and Grubbs elimination at 5 %)

Table II : REPEATABILITY - Absolute difference between replicates in mg / 100 g

Sample Lab code	1	2	3	4	5	6	7	8	9	10	Sr	NL
1	0,86	0,43	0,43	0,86	0,21	0,64	0,86	0,86	0,43	0,64	0,47	20
2	1,00	2,00	1,00	1,00	1,00	1,00	1,00	0,00	0,00	1,00	0,74	20
3	7,00 *	16,00 *	2,00	2,00	1,00	1,00	4,00	3,00	3,00	1,00	4,18	20
4	1,50	0,40	2,60	0,50	0,80	1,10	2,00	0,20	0,10	0,40	0,88	20
5	1,10	0,70	0,90	1,30	1,00	1,30	0,90	1,00	3,00	2,20	1,06	20
6	0,60	0,20	0,00	0,90	0,10	0,90	1,70	1,20	0,40	0,00	0,57	20
7	0,00	0,43	0,64	2,14	0,64	0,43	0,64	2,14	0,64	0,86	0,77	20
8	1,00	1,10	0,40	1,60	0,50	1,20	0,20	1,10	0,20	2,90	0,90	20
9	3,06	2,02	8,82 *	2,67	1,65	1,81	1,87	4,14	4,00	1,48	2,68	20
10	1,34	1,34	1,65	2,06	1,44	0,21	1,03	2,27	0,21	0,21	0,97	20
11	0,58	0,58	0,39	2,13	0,78	0,10	1,36	0,19	0,48	1,74	0,75	20
12	0,00	0,80	1,40	0,60	1,20	2,50	0,20	1,50	0,40	0,40	0,81	20
13	0,60	0,60	1,20	0,60	0,60	0,60	1,20	0,00	0,00	0,60	0,50	20
14	2,30	0,00	2,00	0,40	0,40	0,00	1,10	0,80	1,30	1,90	0,92	20
15	0,31	0,72	1,75	0,21	1,96	1,24	0,52	1,34	0,72	0,21	0,77	20
Sr	1,56	3,00	1,87	1,04	0,72	0,80	1,08	1,22	1,12	0,93		300
NE	30	30	30	30	30	30	30	30	30	30		
L	3,42	2,62	3,62	3,91	2,71	3,01	4,07	4,60	4,20	3,50		

Sr : repeatability standard deviation of each laboratory limit 0,54 mg/100g

NL : number of measurements per laboratory

L : Limit for difference between duplicates according Cochran test at 5% level.

SE : repeatability standard deviation per sample

NE : number of measurements per sample

*: discarded data using the test of Cochran at 5 %

** : missing data

r : limit of repeatability, absolute difference between two replicates=1,50 according ISO 14637|IDF 195

Table III : Means of the replicates in mg / 100 g

(*) : Data modified in right units.

Sample Lab code	1	2	3	4	5	6	7	8	9	10
(*) 1	34,24	46,22	32,10	22,47	63,02	48,47	29,53	55,00	60,13	42,69
2	33,50	43,00	29,50	20,50	61,50	48,50	25,50	52,00	57,00	38,50
3	37,50	43,00	37,00	28,00	69,50	55,50	32,00	58,50	63,50	44,50
4	30,45	39,70	26,60	58,05 *	58,10	44,35	21,90	49,20	53,95	36,30
5	33,25	41,75	26,25	17,65	59,60	45,35	21,85	50,80	55,20	35,00
6	37,90	47,50	32,60	23,05	65,75	51,75	28,25	56,20	61,20	43,20
(*) 7	40,87	49,65	35,42	25,25	68,80	54,14	30,28	61,20	65,16	46,65
8	32,60	41,75	29,60	20,00	29,25 *	48,30	25,50	52,55	58,00	39,35
9	34,21	41,16	29,31	17,82	60,28	47,00	22,88	51,51	56,70	34,94
10	38,99	47,53	31,72	22,87	67,57	51,91	25,75	57,37	62,21	41,30
11	37,69	47,19	33,72	24,32	63,76	50,05	29,07	54,55	59,35	41,18
12	37,90	47,90	33,40	23,40	66,00	51,45	30,30	57,15	62,00	43,50
13	36,90	44,10	33,00	25,50	58,50	47,70	29,40	50,40	54,60	39,90
14	38,65	46,90	31,80	22,30	66,20	50,30	29,05	55,90	59,95	41,85
15	38,88	47,74	33,11	24,21	67,31	53,46	28,48	57,11	61,75	41,10
M	36,24	45,01	31,68	22,67	63,99	49,88	27,32	54,63	59,38	40,66
REF.	36,24	45,01	31,68	22,67	63,99	49,88	27,32	54,63	59,38	40,66
SD	2,96	3,08	2,98	2,90	3,87	3,20	3,22	3,44	3,38	3,40

M = mean per sample

REF. = reference values

SD = standard deviation per sample

*: discarded data using the test of Grubbs 5 %

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 14 laboratories using the Routine method , after discard outliers with Grubbs test at 5 %.

Table IV : Outlier identification

Sample	1	2	3	4	5	6	7	8	9	10
Outliers Cochran	3	3	9							
Outlier Grubbs				4	8					
sr	0,92	0,71	0,98	1,07	0,74	0,80	1,08	1,22	1,12	0,93
SR	3,12	3,18	3,09	3,00	3,90	3,25	3,31	3,55	3,47	3,47

Table V : ACCURACY - differences (laboratory - reference) in mg / 100 g

Sample Lab Code	1	2	3	4	5	6	7	8	9	10	d	Sd _{lab}	t
1	- 2,00	+ 1,22	+ 0,42	- 0,20	- 0,97	- 1,41	+ 2,22	+ 0,37	+ 0,75	+ 2,03	+ 0,24	1,40	0,55
2	- 2,74	- 2,01	- 2,18	- 2,17	- 2,49	- 1,38	- 1,82	- 2,63	- 2,38	- 2,16	- 2,19	0,40	17,31
3	+ 1,26	- 2,01	+ 5,32	+ 5,33	+ 5,51	+ 5,62	+ 4,68	+ 3,87	+ 4,12	+ 3,84	+ 3,76	2,40	4,94
4	- 5,79	- 5,31	- 5,08	+ 35,38	- 5,89	- 5,53	- 5,42	- 5,43	- 5,43	- 4,36	- 1,28	12,89	0,32
5	- 2,99	- 3,26	- 5,43	- 5,02	- 4,39	- 4,53	- 5,47	- 3,83	- 4,18	- 5,66	- 4,47	0,93	15,19
6	+ 1,66	+ 2,49	+ 0,92	+ 0,38	+ 1,76	+ 1,87	+ 0,93	+ 1,57	+ 1,82	+ 2,54	+ 1,60	0,68	7,38
7	+ 4,64	+ 4,64	+ 3,74	+ 2,59	+ 4,81	+ 4,26	+ 2,97	+ 6,57	+ 5,78	+ 5,99	+ 4,60	1,28	11,32
8	- 3,64	- 3,26	- 2,08	- 2,67	- 34,74	- 1,58	- 1,82	- 2,08	- 1,38	- 1,31	- 5,45	10,32	1,67
9	- 2,03	- 3,85	- 2,37	- 4,85	- 3,72	- 2,89	- 4,44	- 3,12	- 2,68	- 5,72	- 3,57	1,18	9,59
10	+ 2,75	+ 2,53	+ 0,05	+ 0,20	+ 3,58	+ 2,03	- 1,57	+ 2,74	+ 2,83	+ 0,64	+ 1,58	1,65	3,03
11	+ 1,46	+ 2,18	+ 2,05	+ 1,66	- 0,23	+ 0,17	+ 1,75	- 0,08	- 0,03	+ 0,52	+ 0,94	0,96	3,10
12	+ 1,66	+ 2,89	+ 1,72	+ 0,73	+ 2,01	+ 1,57	+ 2,98	+ 2,52	+ 2,62	+ 2,84	+ 2,16	0,73	9,28
13	+ 0,66	- 0,91	+ 1,32	+ 2,83	- 5,49	- 2,18	+ 2,08	- 4,23	- 4,78	- 0,76	- 1,14	2,95	1,23
14	+ 2,41	+ 1,89	+ 0,12	- 0,37	+ 2,21	+ 0,42	+ 1,73	+ 1,27	+ 0,57	+ 1,19	+ 1,15	0,93	3,88
15	+ 2,65	+ 2,73	+ 1,44	+ 1,54	+ 3,32	+ 3,58	+ 1,16	+ 2,48	+ 2,37	+ 0,43	+ 2,17	1,00	6,88
d	- 0,00	- 0,00	+ 0,00	+ 0,00	- 0,00	- 0,00	+ 0,00	+ 0,00	- 0,00	+ 0,00	+ 0,00	5,13	
Sd	2,96	3,08	2,98	2,90	3,87	3,20	3,22	3,44	3,38	3,40	3,26		

d = mean of differences

Sd = standard deviation of differences

t = Student test - comparison to 0

Upper limits : $\bar{d} = \pm 2,50 \text{ mg / 100 g}$ Sd = 1,50 mg / 100 g**ISO 14637/IDF 195 : Precision of the method :**

Sr = 0.54 mg / 100 g

SR = 1.81 mg / 100 g

Table VI : Zscore of the different laboratories for each sample.
ZS calculated on the PT standard deviation

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	-0,67	+0,40	+0,14	-0,07	-0,25	-0,44	+0,69	+0,11	+0,22	+0,60
2	-0,92	-0,65	-0,73	-0,75	-0,64	-0,43	-0,56	-0,77	-0,70	-0,64
3	+0,43	-0,65	+1,79	+1,84	+1,42	+1,76	+1,45	+1,13	+1,22	+1,13
4	-1,95	-1,72	-1,70	+12,18	-1,52	-1,73	-1,68	-1,58	-1,61	-1,28
5	-1,01	-1,06	-1,82	-1,73	-1,14	-1,42	-1,70	-1,11	-1,24	-1,66
6	+0,56	+0,81	+0,31	+0,13	+0,45	+0,58	+0,29	+0,46	+0,54	+0,74
7	+1,57	+1,51	+1,26	+0,89	+1,24	+1,33	+0,92	+1,91	+1,71	+1,76
8	-1,23	-1,06	-0,70	-0,92	-8,98	-0,49	-0,56	-0,61	-0,41	-0,39
9	-0,68	-1,25	-0,79	-1,67	-0,96	-0,90	-1,38	-0,91	-0,79	-1,68
10	+0,93	+0,82	+0,02	+0,07	+0,92	+0,63	-0,49	+0,80	+0,84	+0,19
11	+0,49	+0,71	+0,69	+0,57	-0,06	+0,05	+0,54	-0,02	-0,01	+0,15
12	+0,56	+0,94	+0,58	+0,25	+0,52	+0,49	+0,93	+0,73	+0,78	+0,83
13	+0,22	-0,29	+0,44	+0,98	-1,42	-0,68	+0,65	-1,23	-1,42	-0,22
14	+0,82	+0,61	+0,04	-0,13	+0,57	+0,13	+0,54	+0,37	+0,17	+0,35
15	+0,89	+0,89	+0,48	+0,53	+0,86	+1,12	+0,36	+0,72	+0,70	+0,13

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 2 :

Zscore of the different laboratories for each sample. ZS calculated on the PT standard deviation




Table VII : Zscore of the different laboratories for each sample.
ZS calculated on the standard deviation of reproducibility of the method

Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	-1,10	+0,67	+0,23	-0,11	-0,54	-0,78	+1,22	+0,20	+0,42	+1,12
2	-1,51	-1,11	-1,20	-1,20	-1,38	-0,76	-1,00	-1,45	-1,32	-1,20
3	+0,70	-1,11	+2,94	+2,95	+3,04	+3,10	+2,59	+2,14	+2,28	+2,12
4	-3,20	-2,93	-2,80	+19,55	-3,26	-3,06	-2,99	-3,00	-3,00	-2,41
5	-1,65	-1,80	-3,00	-2,77	-2,43	-2,50	-3,02	-2,12	-2,31	-3,13
6	+0,92	+1,38	+0,51	+0,21	+0,97	+1,03	+0,52	+0,87	+1,01	+1,40
7	+2,56	+2,56	+2,07	+1,43	+2,66	+2,35	+1,64	+3,63	+3,19	+3,31
8	-2,01	-1,80	-1,15	-1,47	-19,19	-0,87	-1,00	-1,15	-0,76	-0,73
9	-1,12	-2,13	-1,31	-2,68	-2,05	-1,59	-2,45	-1,72	-1,48	-3,16
10	+1,52	+1,40	+0,03	+0,11	+1,98	+1,12	-0,87	+1,51	+1,56	+0,35
11	+0,81	+1,21	+1,13	+0,91	-0,13	+0,09	+0,97	-0,04	-0,02	+0,29
12	+0,92	+1,60	+0,95	+0,41	+1,11	+0,87	+1,65	+1,39	+1,45	+1,57
13	+0,37	-0,50	+0,73	+1,57	-3,03	-1,21	+1,15	-2,34	-2,64	-0,42
14	+1,33	+1,05	+0,07	-0,20	+1,22	+0,23	+0,96	+0,70	+0,31	+0,65
15	+1,46	+1,51	+0,79	+0,85	+1,83	+1,98	+0,64	+1,37	+1,31	+0,24


This table will allows to compare your ZSCORE from one PT to an other because the standard deviation has always the value of SR of the method SR=1,81

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 3 :

Zscore of the different laboratories for each sample. ZS calculated on the standard deviation of reproducibility of the method






Figure 1 : ACCURACY - Evaluation of the individual performances (to see table I).

LIST OF ICAR PARTICIPANTS

RAW MILK
UREA ROUTINE METHOD
MARCH 2017

CATTLE INFORMATION SERVICE (CIS)
COOPRINSEM
CROATIAN AGRICULTURAL AG., CENTRAL LAB. for MILK QUALITY CONTROL
EASTERN LAB SERVICES
EUROFINS STEINS LABORATORY A/B
FED.LATTERIE SOCIALI DI BOLZANO
LAB. POLJOPRIVREDNI
LOM KCHZ LABORATORIUM REFERENCYJNE z/s w
MERIEUX NUTRISCIENCE
MERIEUX NUTRISCIENCE
MERIEUX NUTRISCIENCE J BAY
MILCHPRUFRING BADEN-WURTTENBERG E.V.
SHANGAI DAIRY BREEDING CENTER Co.Ltd
TAIWAN LIVESTOCK RESEARCH INT.

TELFORD
OSORNO

KRIZEVCI
MEDINA
JÖNGÖPING
BOLZANO
NOVI SAD
PARZNIEWIE
CAPE TOWN
MIDRAND
CAPE TOWN
KIRCHEIM / TECK
SHANGHAI
TAIWAN

UNITED KINGDOM
CHILE

CROATIA
USA
SWEDEN
ITALY
SERBIA
POLAND
SOUTH AFRICA
SOUTH AFRICA
SOUTH AFRICA
GERMANY
CHINA
TAIWAN



ICAR
PROFICIENCY TESTING SCHEME

March 2017

Raw Milk

BHB Beta-HydroxyButyrate

Sending date of statistical treatment : 21th April 2017

Frame of activity :	ICAR Milk Analyses Sub Committee (MA SC)		
Contact :	Gavin Scott	gavin@milkttest.co.nz	
ICAR Staff	Silvia Orlandini	pt@icar.org	silvia@icar.org

Table I : Ranking of the laboratories

Units : milimole of BHB / liter of milk

Nb	%	N°	d	Sd	D	Method
1	6	8	- 0,012	0,020	0,023	A
2	13	15	+ 0,013	0,020	0,024	A
3	19	13	- 0,021	0,015	0,026	A
4	25	7	- 0,022	0,018	0,029	A
5	31	5	+ 0,033	0,020	0,038	A
6	38	3	- 0,036	0,013	0,038	A
7	44	1	+ 0,028	0,026	0,039	B
8	50	14	- 0,045	0,013	0,047	A
9	56	12	- 0,010	0,047	0,048	A
10	63	9	+ 0,056	0,032	0,065	A
11	69	10	+ 0,065	0,042	0,077	A
12	75	6	- 0,081	0,015	0,082	A
13	81	2	- 0,046	0,097	0,107	A
14	88	11	+ 0,006	0,114	0,115	A
15	94	16	+ 0,111	0,062	0,127	A
16	100	4	+ 0,074	0,134	0,153	B

The table should be studied in parallel with figure 1 where the laboratories are located according to an acceptability area (or target) the limits of which are :

—
d = +/- 0,045 milimole of BHB / liter of milk
Sd = 0,045 milimole of BHB / liter of milk

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 16 set of results send by 16 laboratories discarding using Grubbs test at 5 % risk level

A I.R.
B Continuous flow analyzer

(NC : OUT of RANKING because of insufficient data number)

(Nb : laboratory rank; % : relative rank)

(N° : laboratory identification number)

(d et Sd : mean and standard deviation of the differences (laboratory -reference))

(D : Euclidian distance to YX-axis origin = SQUARE ROOT.(d² + Sd²))

Repeatability standard deviation of this ICAR proficiency test (after Cochran elimination at 5 %)

Sr_{PT} 0,011

Reproducibility standard deviation of this ICAR proficiency test (after Cochran and Grubbs elimination at 5 %)

SR_{PT} 0,058

Note : Limits are only indicative and so far do not constitute standard values; they indicate what is normally reachable by labs for their self evaluation.

Table II : REPEATABILITY - Absolute difference between replicates in milimole of BHB / liter of milk

Sample Lab code	1	2	3	4	5	6	7	8	9	10	Sr	NL
1	0,000	0,000	0,000	0,010	0,000	0,000	0,010	0,010	0,000	0,000	0,004	20
2	0,010	0,010	0,020	0,010	0,010	0,010	0,030	0,010	0,010	0,000	0,010	20
3	0,009	0,008	0,013	0,012	0,007	0,009	0,044	0,011	0,014	0,001	0,012	20
4	0,006	0,000	0,002	0,001	0,000	0,002	0,001	0,002	0,001	0,002	0,002	20
5	0,000	0,000	0,010	0,000	0,010	0,000	0,000	0,000	0,000	0,000	0,003	20
6	**	0,010	0,010	0,020	0,000	0,010	0,010	**	0,000	0,010	0,008	16
7	0,000	0,000	0,000	0,010	0,000	0,010	0,050	0,010	0,010	0,010	0,012	20
8	0,000	0,010	0,010	0,010	0,010	0,010	0,010	0,000	0,010	0,010	0,006	20
9	0,010	0,020	0,030	0,010	0,050 *	0,020	0,060	0,010	0,010	0,010	0,020	20
10	0,010	0,070 *	0,050	0,020	0,080 *	0,000	0,040	0,030	0,000	0,020	0,029	20
11	0,010	0,000	0,020	0,030	0,000	0,000	0,010	0,020	0,020	0,000	0,011	20
12	**	**	**	**	**	**	**	**	**	**		
13	0,010	0,030	0,030	0,020	0,020	0,010	0,030	0,030	0,010	0,010	0,015	20
14	0,000	0,020	0,010	0,010	0,000	0,010	0,020	0,010	0,010	0,020	0,009	20
15	0,000	0,010	0,000	0,010	0,010	0,010	0,020	0,000	0,010	0,010	0,007	20
16	**	**	**	**	**	**	**	**	**	**		
Sr	0,005	0,016	0,014	0,010	0,019	0,007	0,021	0,010	0,007	0,007		276
NE	26	28	28	28	28	28	28	26	28	28		
L	0,018	0,033	0,053	0,038	0,021	0,024	0,079	0,038	0,025	0,026		

Sr : repeatability standard deviation of each laboratory limit 0,011 milimole of BHB / liter of milk

NL : number of measurements per laboratory

L : Limit for difference between duplicates according Cochran test at 5% level.

SE : repeatability standard deviation per sample

NE : number of measurements per sample

*: discarded data using the test of Cochran at 5 %

** : missing data

r : limit of repeatability, absolute difference between two replicates=0,030 milimole of BHB / liter of milk

Table III : Means of the replicates in milimole of BHB / liter of milk

Sample Lab Code	1	2	3	4	5	6	7	8	9	10
1	0,050	0,230	0,100	0,125	0,280	0,180	0,205	0,075	0,250	0,150
2	0,145	0,045	0,050	0,185	0,085	0,115	0,025	0,175	0,065	0,010
3	0,026	0,150	0,040	0,075	0,175	0,111	0,137	0,031	0,184	0,075
4	0,305 *	0,124	0,163	0,375 *	0,224	0,260	0,083	0,330 *	0,184	0,055
5	0,080	0,230	0,115	0,130	0,285	0,160	0,200	0,090	0,250	0,150
6		0,085	0,015	0,020	0,140	0,055	0,075	0,010	0,120	0,045
7	0,020	0,160	0,070	0,065	0,220	0,125	0,145	0,035	0,205	0,095
8	0,030	0,195	0,095	0,085	0,215	0,115	0,145	0,050	0,205	0,105
9	0,085	0,250	0,135	0,135	0,335	0,180	0,230	0,105	0,295	0,175
10	0,065	0,305	0,135	0,160	0,360	0,210	0,230	0,115	0,280	0,150
11	0,215	0,060	0,090	0,265	0,140	0,190	0,055	0,250 *	0,130	0,030
12	0,140	0,150	0,070	0,100	0,170					
13	0,025	0,175	0,065	0,090	0,190	0,125	0,145	0,045	0,205	0,085
14	0,020	0,140	0,035	0,065	0,190	0,085	0,120	0,025	0,155	0,080
15	0,050	0,205	0,100	0,105	0,255	0,145	0,180	0,080	0,225	0,145
16	0,100	0,340	0,160	0,190	0,450	0,240	0,310	0,110	0,360	0,210
M	0,075	0,178	0,090	0,120	0,232	0,153	0,152	0,073	0,208	0,104
REF.	0,071	0,176	0,090	0,116	0,227	0,152	0,150	0,070	0,207	0,103
SD	0,058	0,082	0,044	0,061	0,093	0,057	0,076	0,046	0,075	0,058

M = mean per sample

REF. = reference values

SD = standard deviation per sample

*: discarded data using the test of Grubbs at 5 %

REF : Assigned values are robust average values per sample according to algorithm A of standard ISO 13528, of 16 laboratories , after outliers discarding using Grubbs test at 5 % risk level.

Table IV : Outlier identification

Sample	1	2	3	4	5	6	7	8	9	10
Outliers Cochran		10			9; 10					
Outlier Grubbs	4			4				4; 11		
sr	0,005	0,009	0,014	0,011	0,006	0,007	0,021	0,010	0,007	0,007
SR	0,059	0,066	0,044	0,063	0,060	0,054	0,066	0,046	0,064	0,052

Table V : ACCURACY - differences (laboratory - reference) in milimole of BHB / liter of milk

Sample Lab code	1	2	3	4	5	6	7	8	9	10	d	Sd _{lab}	t
1	- 0,021	+ 0,054	+ 0,010	+ 0,009	+ 0,053	+ 0,028	+ 0,055	+ 0,005	+ 0,043	+ 0,047	+ 0,028	0,026	3,42
2	+ 0,074	- 0,131	- 0,040	+ 0,069	- 0,142	- 0,037	- 0,125	+ 0,105	- 0,142	- 0,093	- 0,046	0,097	1,50
3	- 0,045	- 0,026	- 0,050	- 0,041	- 0,053	- 0,042	- 0,013	- 0,039	- 0,023	- 0,029	- 0,036	0,013	8,84
4	+ 0,234	- 0,052	+ 0,073	+ 0,258	- 0,003	+ 0,108	- 0,067	+ 0,260	- 0,023	- 0,048	+ 0,074	0,134	1,75
5	+ 0,009	+ 0,054	+ 0,025	+ 0,014	+ 0,058	+ 0,008	+ 0,050	+ 0,020	+ 0,043	+ 0,047	+ 0,033	0,020	5,30
6	- 0,091	- 0,075	- 0,096	- 0,087	- 0,097	- 0,075	- 0,060	- 0,087	- 0,058	- 0,081	- 0,015	0,015	16,61
7	- 0,051	- 0,016	- 0,020	- 0,051	- 0,007	- 0,027	- 0,005	- 0,035	- 0,002	- 0,008	- 0,022	0,018	3,83
8	- 0,041	+ 0,019	+ 0,005	- 0,031	- 0,012	- 0,037	- 0,005	- 0,020	- 0,002	+ 0,002	- 0,012	0,020	1,95
9	+ 0,014	+ 0,074	+ 0,045	+ 0,019	+ 0,108	+ 0,028	+ 0,080	+ 0,035	+ 0,088	+ 0,072	+ 0,056	0,032	5,52
10	- 0,006	+ 0,129	+ 0,045	+ 0,044	+ 0,133	+ 0,058	+ 0,080	+ 0,045	+ 0,073	+ 0,047	+ 0,065	0,042	4,92
11	+ 0,144	- 0,116	+ 0,000	+ 0,149	- 0,087	+ 0,038	- 0,095	+ 0,180	- 0,077	- 0,073	+ 0,006	0,114	0,18
12	+ 0,069	- 0,026	- 0,020	- 0,016	- 0,057						- 0,010	0,047	0,47
13	- 0,046	- 0,001	- 0,025	- 0,026	- 0,037	- 0,027	- 0,005	- 0,025	- 0,002	- 0,018	- 0,021	0,015	4,49
14	- 0,051	- 0,036	- 0,055	- 0,051	- 0,037	- 0,067	- 0,030	- 0,045	- 0,052	- 0,023	- 0,045	0,013	10,73
15	- 0,021	+ 0,029	+ 0,010	- 0,011	+ 0,028	- 0,007	+ 0,030	+ 0,010	+ 0,018	+ 0,042	+ 0,013	0,020	1,98
16	+ 0,029	+ 0,164	+ 0,070	+ 0,074	+ 0,223	+ 0,088	+ 0,160	+ 0,040	+ 0,153	+ 0,107	+ 0,111	0,062	5,63
d	+ 0,004	+ 0,002	- 0,000	+ 0,004	+ 0,005	+ 0,001	+ 0,002	+ 0,003	+ 0,001	+ 0,001	+ 0,008	0,074	
Sd	0,058	0,082	0,044	0,061	0,093	0,057	0,076	0,046	0,075	0,058	0,067		

d = mean of differences

Sd = standard deviation of differences

t = Student test - comparison to 0

Upper limits : $\bar{d} = \pm 0,045$ milimole of BHB / liter of milk
Sd = 0,045 milimole of BHB / liter of milk

Table VI : Zscore of the different laboratories for each sample.
ZS calculated on the PT standard deviation


Sample Lab code	1	2	3	4	5	6	7	8	9	10
1	-0,35	+0,66	+0,23	+0,14	+0,57	+0,48	+0,72	+0,11	+0,58	+0,81
2	+1,28	-1,60	-0,90	+1,12	-1,53	-0,65	-1,65	+2,27	-1,90	-1,62
3	-0,78	-0,31	-1,14	-0,67	-0,56	-0,73	-0,17	-0,85	-0,30	-0,50
4	+4,04	-0,63	+1,65	+4,20	-0,03	+1,88	-0,89	+5,62	-0,31	-0,84
5	+0,16	+0,66	+0,57	+0,23	+0,62	+0,13	+0,66	+0,43	+0,58	+0,81
6		-1,11	-1,69	-1,56	-0,94	-1,70	-0,99	-1,29	-1,16	-1,01
7	-0,87	-0,19	-0,45	-0,83	-0,08	-0,48	-0,07	-0,75	-0,02	-0,14
8	-0,70	+0,23	+0,12	-0,51	-0,13	-0,65	-0,07	-0,43	-0,02	+0,03
9	+0,25	+0,91	+1,02	+0,31	+1,16	+0,48	+1,05	+0,76	+1,18	+1,25
10	-0,10	+1,58	+1,02	+0,71	+1,43	+1,01	+1,05	+0,97	+0,98	+0,81
11	+2,49	-1,41	+0,00	+2,42	-0,94	+0,66	-1,25	+3,89	-1,03	-1,27
12	+1,20	-0,31	-0,45	-0,26	-0,61					
13	-0,78	-0,01	-0,56	-0,43	-0,40	-0,48	-0,07	-0,54	-0,02	-0,32
14	-0,87	-0,44	-1,24	-0,83	-0,40	-1,18	-0,39	-0,97	-0,69	-0,40
15	-0,35	+0,36	+0,23	-0,18	+0,30	-0,13	+0,40	+0,22	+0,24	+0,73
16	+0,51	+2,01	+1,59	+1,20	+2,40	+1,53	+2,11	+0,87	+2,05	+1,86

In yellow the values bigger or smaller than 2/-2

In red the values bigger or smaller than 3/-3

Figure 2 :

Zscore of the different laboratories for each sample. ZS calculated on the PT standard deviation






Figure 1 : ACCURACY - Evaluation of the individual performances (to see table I).

LIST OF ICAR PARTICIPANTS

RAW MILK
BHB
MARCH 2017

Name	City	Country
ACTALIA	POLIGNY	FRANCE
CENTRAL MILK TESTING LAB	ALBERTA	CANADA
ESTONIAN LIVESTOCK PERFORMANCE RECORDING Ltd	TARTU	ESTONIA
EUROFINS STEINS LABORATORY A/S	VEJEN	DENMARK
EUROFINS STEINS LABORATORY A/S	JÖNKÖPING	SWEDEN
HORIZON LAB	WINNIPEG	CANADA
LAB. POLJOPRIVREDNI	NOVI SAD	SERBIA
LABORATORIO REGIONAL DE VETERINARIA	AZORES	PORTUGAL
LIGAL	LA CORUNA	SPAIN
PACIFIC MILK ANALYSIS	CHILLIWACK	CANADA
QLIP N.V.	CM ZUTPHEN	NETHERLAND
TAIWAN LIVESTOCK RESEARCH INST.	HSINHUA	TAIWAN
THE CATTLE INFORMATION SERVICE	TELFORD	UNITED KINGDOM
VALACTA	STE ANNE DE BELLEVUE	CANADA



ICAR
PROFICIENCY TESTING SCHEME

MARCH 2017

Cow Raw Milk

DETECTION of PAG (Pregnancy Associated Glycoproteins)

Sending date of statistical treatment : 21th April 2017

Frame of activity :	ICAR Milk Analyses Sub Committee (MA SC)		
Contact :	Gavin Scott	gavin@milktest.co.nz	
ICAR Staff	Silvia Orlandini	pt@icar.org	silvia@icar.org

Table I: Methods

N°	METHOD USED
1	IDEXX
2	IDEXX
3	IDEXX
4	IDEXX
5	IDEXX
6	IDEXX
7	IDEXX

Table II : Laboratory results

N°	1	2	3	4	5
1	Y	Y	N	Y	N
2	Y	Y	N	Y	N
3	Y	Y	N	Y	N
4	Y	Y	N	Y	N
5	Y	Y	N	Y	N
6	Y	Y	N	Y	N
7	Y	Y	N	Y	N
REF	Y	Y	N	Y	N

Answers : Y = YES; N = NO; to the questions: Presence of PAG (Pregnancy Associated Glycoproteins)

Table III :

SAMPLES	LACTATION PERIOD	Date
1	Pregnant - Artificial insemination	19.12.2016
2	Pregnant - Artificial insemination	11.12.2016
3	Non pregnant	—
4	Pregnant - Artificial insemination	09.10.2016
5	Non pregnant	—

Table IV : Laboratory accuracy with respect to correct results

N°	1	2	3	4	5	FLR%
1	T	T	T	T	T	100
2	T	T	T	T	T	100
3	T	T	T	T	T	100
4	T	T	T	T	T	100
5	T	T	T	T	T	100
6	T	T	T	T	T	100
7	T	T	T	T	T	100
NSR	7	7	7	7	7	
NS	7	7	7	7	7	
FSR%	100	100	100	100	100	


T : True F : False

NSR : number of right answers per sample and criterion

NS : total number of answers per sample and criterion

FSR% : frequency in right answers per sample and criterion

FLR% : relative frequency in right answers per laboratory



List of participant
PAG
March 2017

BIOR INST.of FOOD SAFETY, ANIMAL HEALTH and ENVIROMENT
CROATIAN AGRICULTURAL AG. CENTRAL LAB.for Milk Quality Control
ESTONIAN LIVESTOCK PERFORMANCE RECORDING Ltd
EUROFINS STEINS LABORATORY A/S
LRV-LABORATORIO REGIONAL DE VETERINARIA
QLIP N.V.
THE CATTLE INFORMATION SERVICE